

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR122018\
 Data File : PR034883.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Dec 2018 19:05
 Operator : SM\SJ
 Sample : J6412-05
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 CB512

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 20 04:01:55 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121818CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Dec 18 01:56:32 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.428	3.513	26052625	71078666	13.394	20.390 #
2) SA Decachlor...	10.112	8.411	53591645	119.7E6	27.260	27.221
Target Compounds						
31) L7 AR-1260-1	7.184	6.027	1352.1E6	3639.7E6	14382.788	16931.087
32) L7 AR-1260-2	7.441	6.213	1629.9E6	4228.4E6	14038.222	15538.532
33) L7 AR-1260-3	7.724	6.364	1794.5E6	3875.8E6	12858.782	15613.275
34) L7 AR-1260-4	8.024	6.829	1111.3E6	2654.5E6	12867.932	15524.199
35) L7 AR-1260-5	8.341	7.070	2109.3E6	5925.1E6	11682.645	12250.838

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR122018\
Data File : PR034883.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Dec 2018 19:05
Operator : SM\SJ
Sample : J6412-05
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampled :
CB512

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 20 04:01:55 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121818CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 18 01:56:32 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

