

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR122024\
 Data File : PR069791.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Dec 2024 16:46
 Operator : AJ\MA
 Sample : AR1248ICC200
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 21 02:06:49 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR122024CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Dec 21 02:05:35 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.631	2.961	10660267	61218424	10.246	10.335
2) SA Decachlor...	9.522	8.239	10970626	73061439	21.337	21.210
Target Compounds						
21) L5 AR-1248-1	4.851	4.082	4086118	23015405	215.453	212.536
22) L5 AR-1248-2	5.143	4.331	5847278	33790867	217.317	217.563
23) L5 AR-1248-3	5.354	4.373	6689457	34535566	211.741	219.517
24) L5 AR-1248-4	5.787	4.550	6942105	39320361	213.948	214.069
25) L5 AR-1248-5	5.827	4.964	6928886	38350595	220.722	221.973



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR122024\
Data File : PR069791.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Dec 2024 16:46
Operator : AJ\MA
Sample : AR1248ICC200
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 21 02:06:49 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR122024CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Dec 21 02:05:35 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#21 AR-1248-1


R.T.: 4.851 min
Delta R.T.: 0.000 min
Response: 4086118
Conc: 215.45 ng/ml

Instrument :
ECD_R
ClientSampleId :




#21 AR-1248-1

R.T.: 4.082 min
Delta R.T.: 0.000 min
Response: 23015405
Conc: 212.54 ng/ml




#22 AR-1248-2

R.T.: 5.143 min
Delta R.T.: 0.000 min
Response: 5847278
Conc: 217.32 ng/ml



#22 AR-1248-2


R.T.: 4.331 min
Delta R.T.: 0.000 min
Response: 33790867
Conc: 217.56 ng/ml



#23 AR-1248-3


R.T.: 5.354 min
Delta R.T.: 0.000 min
Response: 6689457
Conc: 211.74 ng/ml

Instrument :
ECD_R
ClientSampleId :




#23 AR-1248-3

R.T.: 4.373 min
Delta R.T.: 0.000 min
Response: 34535566
Conc: 219.52 ng/ml




#24 AR-1248-4

R.T.: 5.787 min
Delta R.T.: 0.000 min
Response: 6942105
Conc: 213.95 ng/ml



#24 AR-1248-4


R.T.: 4.550 min
Delta R.T.: 0.000 min
Response: 39320361
Conc: 214.07 ng/ml



#25 AR-1248-5

R.T.: 5.827 min
Delta R.T.: 0.000 min
Response: 6928886
Conc: 220.72 ng/ml

Instrument :
ECD_R
ClientSampleId :



#25 AR-1248-5

R.T.: 4.964 min
Delta R.T.: 0.000 min
Response: 38350595
Conc: 221.97 ng/ml