

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR122118\
 Data File : PR034934.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Dec 2018 16:48
 Operator : SM\SJ
 Sample : J6431-06
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 A41X2

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 21 00:44:45 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121818CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Dec 18 01:56:32 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.427	3.511	21278175	55241841	10.940	15.847 #
2)	SA Decachlor...	10.111	8.408	36850508	83720937	18.745	19.042
Target Compounds							
31)	L7 AR-1260-1	7.184	6.024	854.5E6	2259.0E6	9089.605	10508.536
32)	L7 AR-1260-2	7.439	6.210	1127.4E6	2982.8E6	9710.565	10961.121
33)	L7 AR-1260-3	7.723	6.361	1398.2E6	2517.2E6	10018.653	10140.282
34)	L7 AR-1260-4	8.023	6.826	782.6E6	1666.6E6	9061.498	9746.924
35)	L7 AR-1260-5	8.341	7.067	1685.6E6	4720.3E6	9335.757	9759.836

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR122118\
Data File : PR034934.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Dec 2018 16:48
Operator : SM\SJ
Sample : J6431-06
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampled :
A41X2

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 21 00:44:45 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121818CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 18 01:56:32 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

