

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR122218\
 Data File : PR034973.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Dec 2018 14:10
 Operator : SM\SJ
 Sample : J6432-05
 Misc :
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 A41Y4

Manual Integrations
 APPROVED

Sohil
 12/26/2018 8:08:57 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 22 00:39:10 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121818CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Dec 18 01:56:32 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.426	3.511	37720036	72505015	19.393	20.799
2) SA Decachlor...	10.107	8.405	68265333	144.8E6	34.724	32.924
Target Compounds						
31) L7 AR-1260-1	7.182	6.022	48191072	114.2E6	512.627	531.274
32) L7 AR-1260-2	7.438	6.208	65129545	158.0E6	560.972	580.736
33) L7 AR-1260-3	7.721	6.358	80719910	131.3E6	578.403m	529.083
34) L7 AR-1260-4	8.020	6.824	50021184	88694762	579.193m	518.712
35) L7 AR-1260-5	8.337	7.064	100.3E6	272.4E6	555.766	563.239

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR122218\
Data File : PR034973.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Dec 2018 14:10
Operator : SM\SJ
Sample : J6432-05
Misc :
ALS Vial : 43 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
A41Y4

Manual Integrations
APPROVED

Sohil
12/26/2018 8:08:57 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 22 00:39:10 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121818CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 18 01:56:32 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

