

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR122218\
Data File : PR035009.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Dec 2018 00:21
Operator : SM\SJ
Sample : J6432-13
Misc :
ALS Vial : 51 Sample Multiplier: 1


Instrument :
ECD_R
ClientSampleId :
A41Z2

Manual Integrations
APPROVED

Sohil
12/26/2018 8:20:33 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 22 03:41:58 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121818CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 18 01:56:32 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(26) AR-1254-1 (L6)
R.T. Response Conc
6.44 33292702 407.47
6.65 21222883 166.09
7.02 27872241 206.50
7.30 13053617 123.06
7.72 39562018 369.24

(26) AR-1254-1 #2 (L6)
R.T. Response Conc
5.35 27347538 111.76
5.50 33339522 156.75
5.90 86255209 241.40
6.12 29469263 124.77
6.53 93066149 291.80

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR122218\
Data File : PR035009.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Dec 2018 00:21
Operator : SM\SJ
Sample : J6432-13
Misc :
ALS Vial : 51 Sample Multiplier: 1


Instrument :
ECD_R
ClientSampleId :
A41Z2

Manual Integrations
APPROVED

Sohil
12/26/2018 8:20:33 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 22 03:41:58 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121818CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 18 01:56:32 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(26) AR-1254-1 #2 (L6)
R.T. Response Conc
6.44 30047346 367.75
6.65 21222883 166.09
7.02 22186958 164.38
7.30 14407293 135.82
7.72 39562018 369.24

(26) AR-1254-1 #2 (L6)
R.T. Response Conc
5.35 27347538 111.76
5.50 33339522 156.75
5.89 64165236 179.57
6.12 29469263 124.77
6.53 93066149 291.80

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR122218\
Data File : PR035009.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Dec 2018 00:21
Operator : SM\SJ
Sample : J6432-13
Misc :
ALS Vial : 51 Sample Multiplier: 1


Instrument :
ECD_R
ClientSampled :
A41Z2

Manual Integrations
APPROVED

Sohil
12/26/2018 8:20:33 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 22 03:41:58 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121818CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 18 01:56:32 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(31) AR-1260-1 (L7)

R.T.	Response	Conc
7.18	35979859	382.73
7.44	31019066	267.17
7.72	39562018	283.48
8.02	23341033	270.26
8.34	34522901	191.21

(31) AR-1260-1 #2 (L7)

R.T.	Response	Conc
6.02	61049728	283.99
6.21	87908647	323.05
6.36	58194663	234.43
6.82	34143070	199.68
7.06	108883221	225.13

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR122218\
Data File : PR035009.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Dec 2018 00:21
Operator : SM\SJ
Sample : J6432-13
Misc :
ALS Vial : 51 Sample Multiplier: 1


Instrument :
ECD_R
ClientSampled :
A41Z2

Manual Integrations
APPROVED

Sohil
12/26/2018 8:20:33 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 22 03:41:58 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121818CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 18 01:56:32 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(31) AR-1260-1 #2 (L7)
R.T. Response Conc
7.18 23922851 254.48
7.44 31019066 267.17
7.72 39562018 283.48
8.02 23341033 270.26
8.34 36185730 200.42
(31) AR-1260-1 #2 (L7)
R.T. Response Conc
6.02 61049728 283.99
6.21 82192575 302.04
6.36 58194663 234.43
6.82 34143070 199.68
7.06 108883221 225.13