

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR122222\
Data File : PR058715.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Dec 2022 06:18
Operator : YP\AJ
Sample : N6104-15
Misc :
ALS Vial : 71 Sample Multiplier: 1


Instrument :
ECD_R
ClientSampleId :
EXGN4

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 12/23/2022
Supervised By :Ankita Jodhani 12/23/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 23 07:59:00 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121422CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 15 03:47:13 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(31) AR-1260-1 (L7)
R.T. Response Conc
6.66 53081582 509.85
6.93 128959932 1041.34
7.33 46022685 490.64
7.58 62895002 583.57
7.92 93052157 447.74

(31) AR-1260-1 #2 (L7)
R.T. Response Conc
5.58 109280038 634.70
5.78 91802468 436.68
5.94 84292983 423.81
6.44 77354141 465.85
6.71 173724800 440.54

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR122222\
 Data File : PR058715.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Dec 2022 06:18
 Operator : YP\AJ
 Sample : N6104-15
 Misc :
 ALS Vial : 71 Sample Multiplier: 1


Instrument :
 ECD_R
 ClientSampleId :
 EXGN4

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 12/23/2022
 Supervised By :Ankita Jodhani 12/23/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 23 07:59:00 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121422CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 15 03:47:13 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(31) AR-1260-1 #2 (L7)
 R.T. Response Conc
 6.66 53081582 509.85
 6.93 128959932 1041.34
 7.33 46022685 490.64
 7.58 52916178 490.99
 7.92 93052157 447.74

(31) AR-1260-1 #2 (L7)
 R.T. Response Conc
 5.58 109280038 634.70
 5.78 91802468 436.68
 5.94 84292983 423.81
 6.44 77354141 465.85
 6.71 173724800 440.54