

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR122319\
Data File : PR044086.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Dec 2019 10:13
Operator : SM\AJ
Sample : K6277-04
Misc :
ALS Vial : 12 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
C0C40

Manual Integrations
APPROVED

mohammad
12/31/2019 9:11:31 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 24 02:44:31 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121519CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Dec 16 06:15:17 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(31) AR-1260-1 (L7)		
R.T.	Response	Conc
7.48	22217151	96.11
7.73	28843119	101.09
8.09	21702617	96.65
8.33	27092626	104.63
8.67	48155168	86.21

(31) AR-1260-1 #2 (L7)		
R.T.	Response	Conc
6.54	68561164	125.46
6.72	111976149	139.53
6.88	96350273	144.27
7.35	69578440	123.95
7.59	150304723	89.29

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR122319\
Data File : PR044086.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Dec 2019 10:13
Operator : SM\AJ
Sample : K6277-04
Misc :
ALS Vial : 12 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
C0C40

Manual Integrations
APPROVED

mohammad
12/31/2019 9:11:31 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 24 02:44:31 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121519CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Dec 16 06:15:17 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(31) AR-1260-1 (L7)
R.T. Response Conc
7.48 22217151 96.11
7.73 28843119 101.09
8.09 21702617 96.65
8.33 27092626 104.63
8.67 48155168 86.21

(31) AR-1260-1 #2 (L7)
R.T. Response Conc
6.54 41115633 75.24
6.72 80190225 99.92
6.88 56433988 84.50
7.35 53815540 95.87
7.59 141684634 84.17

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR122319\
 Data File : PR044086.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Dec 2019 10:13
 Operator : SM\AJ
 Sample : K6277-04
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 C0C40

Manual Integrations
 APPROVED

mohammad
 12/31/2019 9:11:31 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 24 02:44:31 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121519CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Dec 16 06:15:17 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.695	3.955	43698383	111.6E6	10.933	10.108
2) SA Decachlor...	10.592	9.038	49328087	106.3E6	10.338	10.539
Target Compounds						
31) L7 AR-1260-1	7.477	6.537	22217151	41115633	96.109	75.235m
32) L7 AR-1260-2	7.732	6.722	28843119	80190225	101.087	99.921m
33) L7 AR-1260-3	8.093	6.878	21702617	56433988	96.653	84.500m
34) L7 AR-1260-4	8.330	7.350	27092626	53815540	104.630	95.866m
35) L7 AR-1260-5	8.667	7.590	48155168	141.7E6	86.213	84.167m

AJ
 12/30/19

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.