

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR122324\
 Data File : PR069823.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Dec 2024 10:25
 Operator : AJ\MA
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 24 00:23:22 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR122024CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Dec 21 03:09:04 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
2)	SA Decachlor...	0.000	8.244	0	642578	N.D.	0.191 #
Target Compounds							
7)	L1 AR-1016-5	0.000	4.493f	0	74104	N.D.	0.551 #
15)	L3 AR-1232-5	0.000	4.493f	0	74104	N.D.	1.342 #
20)	L4 AR-1242-5	0.000	4.948	0	1984246	N.D.	14.906 #
24)	L5 AR-1248-4	0.000	4.493f	0	74104	N.D.	0.403 #
25)	L5 AR-1248-5	0.000	4.948	0	1984246	N.D.	11.284 #
26)	L6 AR-1254-1	0.000	4.948	0	1984246	N.D.	7.523 #
27)	L6 AR-1254-2	0.000	5.107	0	428721	N.D.	1.885 #
28)	L6 AR-1254-3	0.000	5.474	0	52993	N.D.	0.152 #
29)	L6 AR-1254-4	0.000	5.753	0	475470	N.D.	2.298 #
30)	L6 AR-1254-5	0.000	6.155f	0	184076	N.D.	0.622 #
31)	L7 AR-1260-1	0.000	5.715f	0	3530893	N.D.	15.012 #
32)	L7 AR-1260-2	0.000	5.814f	0	979473	N.D.	3.649 #
33)	L7 AR-1260-3	0.000	5.991	0	1848217	N.D.	7.235 #
34)	L7 AR-1260-4	0.000	6.526	0	2567123	N.D.	11.818 #
35)	L7 AR-1260-5	0.000	6.764	0	1910905	N.D.	3.825 #
36)	L8 AR-1262-1	0.000	6.259	0	3690439	N.D.	11.634 #
37)	L8 AR-1262-2	0.000	6.764	0	1910905	N.D.	3.443 #
38)	L8 AR-1262-3	0.000	7.082	0	1534063	N.D.	6.956 #
40)	L8 AR-1262-5	0.000	7.704	0	268509	N.D.	1.421 #
41)	L9 AR-1268-1	0.000	7.082	0	1534063	N.D.	2.321 #
43)	L9 AR-1268-3	0.000	7.368	0	919790	N.D.	1.769 #
44)	L9 AR-1268-4	0.000	7.704	0	268509	N.D.	1.253 #
45)	L9 AR-1268-5	0.000	7.929f	0	1035765	N.D.	0.641 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR122324\
Data File : PR069823.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Dec 2024 10:25
Operator : AJ\MA
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 24 00:23:22 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR122024CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Dec 21 03:09:04 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#7 AR-1016-5


R.T.: 0.000 min
Exp R.T.: 5.354 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#7 AR-1016-5

R.T.: 4.493 min
Delta R.T.: -0.057 min
Response: 74104
Conc: 0.55 ng/ml




#15 AR-1232-5

R.T.: 0.000 min
Exp R.T.: 5.143 min
Response: 0
Conc: N.D.



#15 AR-1232-5


R.T.: 4.493 min
Delta R.T.: -0.056 min
Response: 74104
Conc: 1.34 ng/ml



#20 AR-1242-5


R.T.: 0.000 min
Exp R.T.: 5.826 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#20 AR-1242-5

R.T.: 4.948 min
Delta R.T.: 0.026 min
Response: 1984246
Conc: 14.91 ng/ml




#24 AR-1248-4

R.T.: 0.000 min
Exp R.T.: 5.788 min
Response: 0
Conc: N.D.



#24 AR-1248-4


R.T.: 4.493 min
Delta R.T.: -0.057 min
Response: 74104
Conc: 0.40 ng/ml



#25 AR-1248-5


R.T.: 0.000 min
Exp R.T.: 5.827 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#25 AR-1248-5

R.T.: 4.948 min
Delta R.T.: -0.015 min
Response: 1984246
Conc: 11.28 ng/ml




#26 AR-1254-1

R.T.: 0.000 min
Exp R.T.: 5.758 min
Response: 0
Conc: N.D.



#26 AR-1254-1


R.T.: 4.948 min
Delta R.T.: 0.027 min
Response: 1984246
Conc: 7.52 ng/ml



#27 AR-1254-2


R.T.: 0.000 min
Exp R.T.: 5.995 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#27 AR-1254-2

R.T.: 5.107 min
Delta R.T.: 0.025 min
Response: 428721
Conc: 1.89 ng/ml




#28 AR-1254-3

R.T.: 0.000 min
Exp R.T.: 6.389 min
Response: 0
Conc: N.D.



#28 AR-1254-3


R.T.: 5.474 min
Delta R.T.: -0.031 min
Response: 52993
Conc: 0.15 ng/ml



#29 AR-1254-4


R.T.: 0.000 min
Exp R.T.: 6.700 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#29 AR-1254-4

R.T.: 5.753 min
Delta R.T.: 0.000 min
Response: 475470
Conc: 2.30 ng/ml




#30 AR-1254-5

R.T.: 0.000 min
Exp R.T.: 7.156 min
Response: 0
Conc: N.D.



#30 AR-1254-5


R.T.: 6.155 min
Delta R.T.: -0.044 min
Response: 184076
Conc: 0.62 ng/ml



#31 AR-1260-1


R.T.: 0.000 min
Exp R.T.: 6.566 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#31 AR-1260-1

R.T.: 5.715 min
Delta R.T.: 0.069 min
Response: 3530893
Conc: 15.01 ng/ml




#32 AR-1260-2

R.T.: 0.000 min
Exp R.T.: 6.848 min
Response: 0
Conc: N.D.



#32 AR-1260-2


R.T.: 5.814 min
Delta R.T.: -0.039 min
Response: 979473
Conc: 3.65 ng/ml



#33 AR-1260-3


R.T.: 0.000 min
Exp R.T.: 7.238 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#33 AR-1260-3

R.T.: 5.991 min
Delta R.T.: -0.020 min
Response: 1848217
Conc: 7.23 ng/ml




#34 AR-1260-4

R.T.: 0.000 min
Exp R.T.: 7.481 min
Response: 0
Conc: N.D.



#34 AR-1260-4


R.T.: 6.526 min
Delta R.T.: 0.008 min
Response: 2567123
Conc: 11.82 ng/ml



#35 AR-1260-5


R.T.: 0.000 min
Exp R.T.: 7.822 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#35 AR-1260-5

R.T.: 6.764 min
Delta R.T.: -0.021 min
Response: 1910905
Conc: 3.83 ng/ml




#36 AR-1262-1

R.T.: 0.000 min
Exp R.T.: 7.240 min
Response: 0
Conc: N.D.



#36 AR-1262-1


R.T.: 6.259 min
Delta R.T.: 0.016 min
Response: 3690439
Conc: 11.63 ng/ml



#37 AR-1262-2


R.T.: 0.000 min
Exp R.T.: 7.823 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#37 AR-1262-2

R.T.: 6.764 min
Delta R.T.: -0.021 min
Response: 1910905
Conc: 3.44 ng/ml




#38 AR-1262-3

R.T.: 0.000 min
Exp R.T.: 8.128 min
Response: 0
Conc: N.D.



#38 AR-1262-3


R.T.: 7.082 min
Delta R.T.: -0.008 min
Response: 1534063
Conc: 6.96 ng/ml



#40 AR-1262-5


R.T.: 0.000 min
Exp R.T.: 8.835 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#40 AR-1262-5

R.T.: 7.704 min
Delta R.T.: 0.006 min
Response: 268509
Conc: 1.42 ng/ml




#41 AR-1268-1

R.T.: 0.000 min
Exp R.T.: 8.124 min
Response: 0
Conc: N.D.



#41 AR-1268-1


R.T.: 7.082 min
Delta R.T.: -0.007 min
Response: 1534063
Conc: 2.32 ng/ml



#43 AR-1268-3


R.T.: 0.000 min
Exp R.T.: 8.431 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#43 AR-1268-3

R.T.: 7.368 min
Delta R.T.: -0.013 min
Response: 919790
Conc: 1.77 ng/ml




#44 AR-1268-4

R.T.: 0.000 min
Exp R.T.: 8.834 min
Response: 0
Conc: N.D.



#44 AR-1268-4


R.T.: 7.704 min
Delta R.T.: 0.007 min
Response: 268509
Conc: 1.25 ng/ml



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T.: 9.216 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :



#45 AR-1268-5

R.T.: 7.929 min
Delta R.T.: -0.063 min
Response: 1035765
Conc: 0.64 ng/ml