

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR122324\
 Data File : PR069836.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Dec 2024 14:53
 Operator : AJ\MA
 Sample : P5368-02
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 YE625

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/26/2024
 Supervised By :Ankita Jodhani 12/27/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 24 00:26:28 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR122024CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Dec 21 03:09:04 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.630	2.961	13415381	65778648	12.910	11.718
2) SA Decachlor...	9.520	8.238	5955291	30770483	11.545	9.169
Target Compounds						
26) L6 AR-1254-1	5.757	4.921	1915650	10524787	54.349	39.903 #
27) L6 AR-1254-2	5.995	5.081	2137646	10181305	39.794	44.767
28) L6 AR-1254-3	6.387	5.524	1636732	13051290	29.872	37.357m#
29) L6 AR-1254-4	6.695	5.751	767926	4482225	21.021	21.665m
30) L6 AR-1254-5	7.151	6.198	5560344	32277971	135.562m	109.065
31) L7 AR-1260-1	6.566	5.644	4748215	22576857	118.223	95.985m
32) L7 AR-1260-2	6.848	5.851	5950108	25223522	133.171	93.981m#
33) L7 AR-1260-3	7.240	6.009	3455050	23803509	97.565	93.175
34) L7 AR-1260-4	7.484	6.517	3936895	18275549	100.923	84.131m
35) L7 AR-1260-5	7.824	6.783	7262435	40004979	103.252	80.080m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR122324\
Data File : PR069836.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Dec 2024 14:53
Operator : AJ\MA
Sample : P5368-02
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
YE625

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/26/2024
Supervised By :Ankita Jodhani 12/27/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 24 00:26:28 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR122024CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Dec 21 03:09:04 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

