

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR122723\
 Data File : PR064828.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Dec 2023 17:46
 Operator : AJ\MA
 Sample : AR1660CCC400
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR16603283

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/28/2023
 Supervised By :Ankita Jodhani 12/28/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 27 20:54:15 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121423CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 15 04:19:57 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.702	3.007	125.1E6	110.5E6	20.701	20.457
2) SA Decachlor...	9.722	8.383	94590946	105.1E6	42.926	40.132m
Target Compounds						
3) L1 AR-1016-1	4.940	4.154	74121153	73748757	409.440	397.436
4) L1 AR-1016-2	4.963	4.172	109.8E6	102.1E6	413.759	402.182
5) L1 AR-1016-3	5.027	4.356	72101399	55065485	427.682	399.059
6) L1 AR-1016-4	5.131	4.407	57388215	39253239	421.464	368.767
7) L1 AR-1016-5	5.450	4.631	55947914	54823127	407.618	391.303
31) L7 AR-1260-1	6.677	5.746	102.1E6	112.4E6	408.638	388.834
32) L7 AR-1260-2	6.963	5.954	116.8E6	135.0E6	408.507	388.500
33) L7 AR-1260-3	7.357	6.116	85193493	129.2E6	401.797	392.445
34) L7 AR-1260-4	7.602	6.631	96379925	105.1E6	419.313	389.491
35) L7 AR-1260-5	7.943	6.899	177.0E6	242.9E6	412.561	388.631

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR122723\
Data File : PR064828.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Dec 2023 17:46
Operator : AJ\MA
Sample : AR1660CCC400
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR16603283

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/28/2023
Supervised By :Ankita Jodhani 12/28/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 27 20:54:15 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121423CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 15 04:19:57 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

