

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR122918\
 Data File : PR035134.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Dec 2018 03:39
 Operator : SM\SJ
 Sample : J6532-08
 Misc :
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 29 09:53:15 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121818CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Dec 18 01:56:32 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	0.000	3.466f	0	93672	N.D.	0.027 #
Target Compounds							
3)	L1 AR-1016-1	0.000	4.578	0	123331	N.D.	0.948 #
4)	L1 AR-1016-2	0.000	4.578	0	123331	N.D.	0.624 #
5)	L1 AR-1016-3	0.000	4.806f	0	467643	N.D.	4.865 #
6)	L1 AR-1016-4	0.000	4.806	0	467643	N.D.	6.190 #
7)	L1 AR-1016-5	0.000	5.051f	0	1075909	N.D.	10.463 #
12)	L3 AR-1232-2	0.000	4.578	0	123331	N.D.	1.194 #
13)	L3 AR-1232-3	0.000	4.806f	0	467643	N.D.	9.682 #
14)	L3 AR-1232-4	0.000	4.806f	0	467643	N.D.	10.444 #
15)	L3 AR-1232-5	0.000	5.051f	0	1075909	N.D.	21.647 #
16)	L4 AR-1242-1	0.000	4.578	0	123331	N.D.	1.086 #
17)	L4 AR-1242-2	0.000	4.578	0	123331	N.D.	0.722 #
18)	L4 AR-1242-3	0.000	4.806f	0	467643	N.D.	5.519 #
19)	L4 AR-1242-4	0.000	4.806f	0	467643	N.D.	5.656 #
20)	L4 AR-1242-5	0.000	5.377	0	351779	N.D.	3.163 #
21)	L5 AR-1248-1	0.000	4.578	0	123331	N.D.	1.265 #
22)	L5 AR-1248-2	0.000	4.806	0	467643	N.D.	3.651 #
23)	L5 AR-1248-3	0.000	4.806f	0	467643	N.D.	3.543 #
24)	L5 AR-1248-4	0.000	5.051f	0	1075909	N.D.	6.539 #
25)	L5 AR-1248-5	0.000	5.400	0	129605	N.D.	0.774 #
26)	L6 AR-1254-1	0.000	5.377	0	351779	N.D.	1.438 #
27)	L6 AR-1254-2	0.000	5.489	0	894343	N.D.	4.205 #
28)	L6 AR-1254-3	0.000	5.896	0	3459282	N.D.	9.681 #
29)	L6 AR-1254-4	0.000	6.118	0	111172	N.D.	0.471 #
30)	L6 AR-1254-5	0.000	6.526	0	574254	N.D.	1.800 #
31)	L7 AR-1260-1	0.000	6.020	0	274020	N.D.	1.275 #
32)	L7 AR-1260-2	0.000	6.197	0	617421	N.D.	2.269 #
33)	L7 AR-1260-3	0.000	6.353	0	211490	N.D.	0.852 #
34)	L7 AR-1260-4	0.000	6.818	0	628907	N.D.	3.678 #
35)	L7 AR-1260-5	0.000	7.058	0	937854	N.D.	1.939 #
36)	L8 AR-1262-1	0.000	6.562	0	298279	N.D.	0.768 #
37)	L8 AR-1262-2	0.000	7.058	0	937854	N.D.	1.318 #
38)	L8 AR-1262-3	0.000	7.401f	0	618677	N.D.	2.025 #
39)	L8 AR-1262-4	0.000	7.401	0	618677	N.D.	1.182 #
40)	L8 AR-1262-5	0.000	7.930	0	96176	N.D.	0.393 #
41)	L9 AR-1268-1	0.000	7.401f	0	618677	N.D.	0.680 #
42)	L9 AR-1268-2	0.000	7.401	0	618677	N.D.	0.737 #
43)	L9 AR-1268-3	0.000	7.612	0	309128	N.D.	0.422 #
44)	L9 AR-1268-4	0.000	7.930	0	96176	N.D.	0.331 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR122918\
Data File : PR035134.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Dec 2018 03:39
Operator : SM\SJ
Sample : J6532-08
Misc :
ALS Vial : 56 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 29 09:53:15 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121818CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 18 01:56:32 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.						


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR122918\
Data File : PR035134.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Dec 2018 03:39
Operator : SM\SJ
Sample : J6532-08
Misc :
ALS Vial : 56 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 29 09:53:15 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121818CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 18 01:56:32 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#3 AR-1016-1


R.T.: 0.000 min
Exp R.T.: 5.590 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#3 AR-1016-1

R.T.: 4.578 min
Delta R.T.: 0.004 min
Response: 123331
Conc: 0.95 ng/ml




#4 AR-1016-2

R.T.: 0.000 min
Exp R.T.: 5.612 min
Response: 0
Conc: N.D.



#4 AR-1016-2


R.T.: 4.578 min
Delta R.T.: -0.014 min
Response: 123331
Conc: 0.62 ng/ml



#5 AR-1016-3


R.T.: 0.000 min
Exp R.T. : 5.675 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#5 AR-1016-3

R.T.: 4.806 min
Delta R.T.: 0.042 min
Response: 467643
Conc: 4.87 ng/ml




#6 AR-1016-4

R.T.: 0.000 min
Exp R.T. : 5.773 min
Response: 0
Conc: N.D.



#6 AR-1016-4


R.T.: 4.806 min
Delta R.T.: 0.000 min
Response: 467643
Conc: 6.19 ng/ml



#7 AR-1016-5


R.T.: 0.000 min
Exp R.T. : 6.065 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#7 AR-1016-5

R.T.: 5.051 min
Delta R.T.: 0.037 min
Response: 1075909
Conc: 10.46 ng/ml




#12 AR-1232-2

R.T.: 0.000 min
Exp R.T. : 5.323 min
Response: 0
Conc: N.D.



#12 AR-1232-2


R.T.: 4.578 min
Delta R.T.: -0.013 min
Response: 123331
Conc: 1.19 ng/ml



#13 AR-1232-3


R.T.: 0.000 min
Exp R.T. : 5.612 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#13 AR-1232-3

R.T.: 4.806 min
Delta R.T.: 0.042 min
Response: 467643
Conc: 9.68 ng/ml




#14 AR-1232-4

R.T.: 0.000 min
Exp R.T. : 5.773 min
Response: 0
Conc: N.D.



#14 AR-1232-4


R.T.: 4.806 min
Delta R.T.: -0.040 min
Response: 467643
Conc: 10.44 ng/ml



#15 AR-1232-5


R.T.: 0.000 min
Exp R.T. : 5.861 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#15 AR-1232-5

R.T.: 5.051 min
Delta R.T.: 0.037 min
Response: 1075909
Conc: 21.65 ng/ml




#16 AR-1242-1

R.T.: 0.000 min
Exp R.T. : 5.590 min
Response: 0
Conc: N.D.



#16 AR-1242-1


R.T.: 4.578 min
Delta R.T.: 0.005 min
Response: 123331
Conc: 1.09 ng/ml



#17 AR-1242-2


R.T.: 0.000 min
Exp R.T. : 5.612 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#17 AR-1242-2

R.T.: 4.578 min
Delta R.T.: -0.013 min
Response: 123331
Conc: 0.72 ng/ml




#18 AR-1242-3

R.T.: 0.000 min
Exp R.T. : 5.674 min
Response: 0
Conc: N.D.



#18 AR-1242-3


R.T.: 4.806 min
Delta R.T.: 0.042 min
Response: 467643
Conc: 5.52 ng/ml



#19 AR-1242-4


R.T.: 0.000 min
Exp R.T. : 5.772 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#19 AR-1242-4

R.T.: 4.806 min
Delta R.T.: -0.039 min
Response: 467643
Conc: 5.66 ng/ml




#20 AR-1242-5

R.T.: 0.000 min
Exp R.T. : 6.504 min
Response: 0
Conc: N.D.



#20 AR-1242-5


R.T.: 5.377 min
Delta R.T.: 0.019 min
Response: 351779
Conc: 3.16 ng/ml



#21 AR-1248-1


R.T.: 0.000 min
Exp R.T.: 5.590 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#21 AR-1248-1

R.T.: 4.578 min
Delta R.T.: 0.004 min
Response: 123331
Conc: 1.26 ng/ml




#22 AR-1248-2

R.T.: 0.000 min
Exp R.T.: 5.860 min
Response: 0
Conc: N.D.



#22 AR-1248-2


R.T.: 4.806 min
Delta R.T.: 0.001 min
Response: 467643
Conc: 3.65 ng/ml



#23 AR-1248-3


R.T.: 0.000 min
Exp R.T. : 6.064 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#23 AR-1248-3

R.T.: 4.806 min
Delta R.T.: -0.039 min
Response: 467643
Conc: 3.54 ng/ml




#24 AR-1248-4

R.T.: 0.000 min
Exp R.T. : 6.465 min
Response: 0
Conc: N.D.



#24 AR-1248-4


R.T.: 5.051 min
Delta R.T.: 0.038 min
Response: 1075909
Conc: 6.54 ng/ml



#25 AR-1248-5


R.T.: 0.000 min
Exp R.T.: 6.503 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#25 AR-1248-5

R.T.: 5.400 min
Delta R.T.: 0.002 min
Response: 129605
Conc: 0.77 ng/ml




#26 AR-1254-1

R.T.: 0.000 min
Exp R.T.: 6.438 min
Response: 0
Conc: N.D.



#26 AR-1254-1


R.T.: 5.377 min
Delta R.T.: 0.020 min
Response: 351779
Conc: 1.44 ng/ml



#27 AR-1254-2


R.T.: 0.000 min
Exp R.T.: 6.655 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#27 AR-1254-2

R.T.: 5.489 min
Delta R.T.: -0.013 min
Response: 894343
Conc: 4.20 ng/ml




#28 AR-1254-3

R.T.: 0.000 min
Exp R.T.: 7.021 min
Response: 0
Conc: N.D.



#28 AR-1254-3


R.T.: 5.896 min
Delta R.T.: -0.003 min
Response: 3459282
Conc: 9.68 ng/ml



#29 AR-1254-4


R.T.: 0.000 min
Exp R.T.: 7.305 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#29 AR-1254-4

R.T.: 6.118 min
Delta R.T.: -0.006 min
Response: 111172
Conc: 0.47 ng/ml




#30 AR-1254-5

R.T.: 0.000 min
Exp R.T.: 7.722 min
Response: 0
Conc: N.D.



#30 AR-1254-5


R.T.: 6.526 min
Delta R.T.: -0.010 min
Response: 574254
Conc: 1.80 ng/ml



#31 AR-1260-1


R.T.: 0.000 min
Exp R.T. : 7.183 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#31 AR-1260-1

R.T.: 6.020 min
Delta R.T.: -0.007 min
Response: 274020
Conc: 1.27 ng/ml




#32 AR-1260-2

R.T.: 0.000 min
Exp R.T. : 7.439 min
Response: 0
Conc: N.D.



#32 AR-1260-2


R.T.: 6.197 min
Delta R.T.: -0.017 min
Response: 617421
Conc: 2.27 ng/ml



#33 AR-1260-3


R.T.: 0.000 min
Exp R.T. : 7.722 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#33 AR-1260-3

R.T.: 6.353 min
Delta R.T.: -0.011 min
Response: 211490
Conc: 0.85 ng/ml




#34 AR-1260-4

R.T.: 0.000 min
Exp R.T. : 8.022 min
Response: 0
Conc: N.D.



#34 AR-1260-4


R.T.: 6.818 min
Delta R.T.: -0.013 min
Response: 628907
Conc: 3.68 ng/ml



#35 AR-1260-5


R.T.: 0.000 min
Exp R.T. : 8.339 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#35 AR-1260-5

R.T.: 7.058 min
Delta R.T.: -0.013 min
Response: 937854
Conc: 1.94 ng/ml




#36 AR-1262-1

R.T.: 0.000 min
Exp R.T. : 7.795 min
Response: 0
Conc: N.D.



#36 AR-1262-1


R.T.: 6.562 min
Delta R.T.: -0.011 min
Response: 298279
Conc: 0.77 ng/ml



#37 AR-1262-2


R.T.: 0.000 min
Exp R.T.: 8.338 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#37 AR-1262-2

R.T.: 7.058 min
Delta R.T.: -0.012 min
Response: 937854
Conc: 1.32 ng/ml




#38 AR-1262-3

R.T.: 0.000 min
Exp R.T.: 8.643 min
Response: 0
Conc: N.D.



#38 AR-1262-3


R.T.: 7.401 min
Delta R.T.: 0.052 min
Response: 618677
Conc: 2.02 ng/ml



#39 AR-1262-4


R.T.: 0.000 min
Exp R.T. : 8.732 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#39 AR-1262-4

R.T.: 7.401 min
Delta R.T.: -0.013 min
Response: 618677
Conc: 1.18 ng/ml




#40 AR-1262-5

R.T.: 0.000 min
Exp R.T. : 9.378 min
Response: 0
Conc: N.D.



#40 AR-1262-5


R.T.: 7.930 min
Delta R.T.: 0.026 min
Response: 96176
Conc: 0.39 ng/ml



#41 AR-1268-1


R.T.: 0.000 min
Exp R.T. : 8.641 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#41 AR-1268-1

R.T.: 7.401 min
Delta R.T.: 0.052 min
Response: 618677
Conc: 0.68 ng/ml




#42 AR-1268-2

R.T.: 0.000 min
Exp R.T. : 8.734 min
Response: 0
Conc: N.D.



#42 AR-1268-2


R.T.: 7.401 min
Delta R.T.: -0.013 min
Response: 618677
Conc: 0.74 ng/ml



#43 AR-1268-3


R.T.: 0.000 min
Exp R.T. : 8.957 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#43 AR-1268-3

R.T.: 7.612 min
Delta R.T.: -0.004 min
Response: 309128
Conc: 0.42 ng/ml



#44 AR-1268-4

R.T.: 0.000 min
Exp R.T. : 9.376 min
Response: 0
Conc: N.D.



#44 AR-1268-4

R.T.: 7.930 min
Delta R.T.: 0.026 min
Response: 96176
Conc: 0.33 ng/ml