

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR122918\
 Data File : PR035151.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Dec 2018 08:13
 Operator : SM\SJ
 Sample : J6441-13MSD
 Misc :
 ALS Vial : 73 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 BF023MSD

Manual Integrations
 APPROVED

mohammad
 12/31/2018 11:11:22 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 29 10:01:26 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121818CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Dec 18 01:56:32 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.428	3.514	39409123	87122469	20.261m	24.992
2)	SA Decachlor...	10.102	8.397	60975421	119.9E6	31.016	27.274
Target Compounds							
3)	L1 AR-1016-1	5.589	4.568	33289003	65382065	493.147	502.458
4)	L1 AR-1016-2	5.612	4.586	74198653	116.2E6	749.278	587.837
5)	L1 AR-1016-3	5.675	4.758	38770612	97258438	658.740	1011.886 #
6)	L1 AR-1016-4	5.766	4.799	41903940	305.7E6	883.362m	4046.521 #
7)	L1 AR-1016-5	6.062	5.007	110.7E6	242.3E6	2325.078	2356.797
26)	L6 AR-1254-1	6.437	5.351	304.8E6	765.4E6	3730.780m	3127.870
27)	L6 AR-1254-2	6.653	5.496	508.2E6	820.7E6	3976.865m	3858.511
28)	L6 AR-1254-3	7.019	5.891	551.4E6	1635.4E6	4085.220	4576.903
29)	L6 AR-1254-4	7.302	6.116	383.5E6	788.5E6	3615.700m	3338.460
30)	L6 AR-1254-5	7.718	6.527	528.4E6	1274.9E6	4931.963	3997.404
31)	L7 AR-1260-1	7.180	6.018	234.9E6	721.4E6	2499.048	3355.672m#
32)	L7 AR-1260-2	7.433	6.204	537.0E6	775.0E6	4625.655	2847.905 #
33)	L7 AR-1260-3	7.718	6.353	528.4E6	649.1E6	3786.476	2614.767 #
34)	L7 AR-1260-4	8.010	6.819	244.9E6	216.0E6	2835.324m	1262.972 #
35)	L7 AR-1260-5	8.334	7.059	284.6E6	766.1E6	1576.414m	1584.071

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR122918\
Data File : PR035151.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Dec 2018 08:13
Operator : SM\SJ
Sample : J6441-13MSD
Misc :
ALS Vial : 73 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
BF023MSD

Manual Integrations
APPROVED

mohammad
12/31/2018 11:11:22 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 29 10:01:26 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121818CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 18 01:56:32 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

