

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR022420\
 Data File : PR044773.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Feb 2020 19:53
 Operator : AJ\MA
 Sample : AR1242ICC800
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1242401

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 25 04:28:07 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR022420CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Feb 25 04:26:51 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.688	3.951	382.2E6	647.1E6	36.813	34.917
2) SA Decachlor...	10.575	9.025	614.0E6	1036.1E6	78.774	79.783
Target Compounds						
16) L4 AR-1242-1	5.864	5.042	209.8E6	238.8E6	721.696	676.612
17) L4 AR-1242-2	5.887	5.061	298.5E6	332.8E6	726.798	687.558
18) L4 AR-1242-3	5.950	5.239	175.7E6	190.0E6	721.490	691.911
19) L4 AR-1242-4	6.050	5.322	147.7E6	178.5E6	716.745	691.650
20) L4 AR-1242-5	6.785	5.846	164.6E6	286.0E6	735.122	710.365

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR022420\
Data File : PR044773.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Feb 2020 19:53
Operator : AJ\MA
Sample : AR1242ICC800
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1242401

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 25 04:28:07 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR022420CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 25 04:26:51 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

