

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR030519\
 Data File : PR035894.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Mar 2019 07:26
 Operator : SM\SJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 05 10:23:03 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR030119.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Mar 02 03:49:12 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.414	3.487	34056094	114.3E6	21.429	21.770
2) SA Decachlor...	10.082	8.361	30099323	117.2E6	18.821	17.857
Target Compounds						
9) L2 AR-1221-2	0.000	3.779	0	1623669	N.D.	31.713 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR030519\
Data File : PR035894.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Mar 2019 07:26
Operator : SM\SJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 05 10:23:03 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR030119.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Mar 02 03:49:12 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.414 min
Delta R.T.: 0.000 min
Response: 34056094
Conc: 21.43 ng/ml

Instrument :
ECD_R
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.487 min
Delta R.T.: 0.000 min
Response: 114332273
Conc: 21.77 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.082 min
Delta R.T.: 0.002 min
Response: 30099323
Conc: 18.82 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.361 min
Delta R.T.: -0.001 min
Response: 117210881
Conc: 17.86 ng/ml



#9 AR-1221-2

R.T.: 0.000 min
Exp R.T.: 4.709 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
I.BLK



#9 AR-1221-2

R.T.: 3.779 min
Delta R.T.: 0.000 min
Response: 1623669
Conc: 31.71 ng/ml