

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR030819\
 Data File : PR036005.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Mar 2019 19:26
 Operator : SM\SJ
 Sample : K1807-21
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 TP-18

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 08 21:51:19 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR030119.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 07 10:54:32 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.411	3.484	28543555	92409723	17.960	17.595
2) SA Decachlor...	10.071	8.353	24947917	94711712	15.600	14.429
Target Compounds						
9) L2 AR-1221-2	0.000	3.776	0	1219642	N.D.	23.822 #
27) L6 AR-1254-2	0.000	5.486	0	2311625	N.D.	8.868 #
28) L6 AR-1254-3	0.000	5.878	0	1216226	N.D.	2.742 #
29) L6 AR-1254-4	7.224	0.000	15236914	0	194.561	N.D. #
30) L6 AR-1254-5	0.000	6.523	0	3204582	N.D.	7.655 #
31) L7 AR-1260-1	7.224	0.000	15236914	0	156.914	N.D. #
32) L7 AR-1260-2	0.000	6.162	0	3934442	N.D.	7.532 #
36) L8 AR-1262-1	0.000	6.523	0	3204582	N.D.	15.911 #
45) L9 AR-1268-5	0.000	8.125	0	1464710	N.D.	0.550 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR030819\
Data File : PR036005.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Mar 2019 19:26
Operator : SM\SJ
Sample : K1807-21
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampled :
TP-18

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 08 21:51:19 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR030119.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 07 10:54:32 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.411 min
Delta R.T.: 0.000 min
Response: 28543555
Conc: 17.96 ng/ml

Instrument :
ECD_R
ClientSampleId :
TP-18




#1 Tetrachloro-m-xylene

R.T.: 3.484 min
Delta R.T.: -0.002 min
Response: 92409723
Conc: 17.60 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.071 min
Delta R.T.: -0.009 min
Response: 24947917
Conc: 15.60 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.353 min
Delta R.T.: -0.010 min
Response: 94711712
Conc: 14.43 ng/ml



#9 AR-1221-2


R.T.: 0.000 min
Exp R.T.: 4.709 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
TP-18




#9 AR-1221-2

R.T.: 3.776 min
Delta R.T.: -0.003 min
Response: 1219642
Conc: 23.82 ng/ml




#27 AR-1254-2

R.T.: 0.000 min
Exp R.T.: 6.641 min
Response: 0
Conc: N.D.



#27 AR-1254-2


R.T.: 5.486 min
Delta R.T.: 0.018 min
Response: 2311625
Conc: 8.87 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T.: 7.007 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
TP-18




#28 AR-1254-3

R.T.: 5.878 min
Delta R.T.: 0.014 min
Response: 1216226
Conc: 2.74 ng/ml




#29 AR-1254-4

R.T.: 7.224 min
Delta R.T.: -0.067 min
Response: 15236914
Conc: 194.56 ng/ml



#29 AR-1254-4


R.T.: 0.000 min
Exp R.T.: 6.089 min
Response: 0
Conc: N.D.



#30 AR-1254-5


R.T.: 0.000 min
Exp R.T.: 7.708 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
TP-18




#30 AR-1254-5

R.T.: 6.523 min
Delta R.T.: 0.025 min
Response: 3204582
Conc: 7.66 ng/ml




#31 AR-1260-1

R.T.: 7.224 min
Delta R.T.: 0.056 min
Response: 15236914
Conc: 156.91 ng/ml



#31 AR-1260-1


R.T.: 0.000 min
Exp R.T.: 5.990 min
Response: 0
Conc: N.D.



#32 AR-1260-2


R.T.: 0.000 min
Exp R.T.: 7.424 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
TP-18




#32 AR-1260-2

R.T.: 6.162 min
Delta R.T.: -0.014 min
Response: 3934442
Conc: 7.53 ng/ml




#36 AR-1262-1

R.T.: 0.000 min
Exp R.T.: 7.780 min
Response: 0
Conc: N.D.



#36 AR-1262-1


R.T.: 6.523 min
Delta R.T.: -0.065 min
Response: 3204582
Conc: 15.91 ng/ml



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T.: 9.756 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
TP-18



#45 AR-1268-5

R.T.: 8.125 min
Delta R.T.: -0.009 min
Response: 1464710
Conc: 0.55 ng/ml