

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR031119\
 Data File : PR036058.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Mar 2019 17:41
 Operator : SM\SJ
 Sample : K1807-39
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 TP-20

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 12 03:51:55 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR030119.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 07 10:54:32 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.408	3.481	20556066	66113384	12.934	12.588
2)	SA Decachlor...	10.070	8.354	10598929	38110107	6.627	5.806
Target Compounds							
20)	L4 AR-1242-5	0.000	5.285	0	2822161	N.D.	15.991 #
25)	L5 AR-1248-5	0.000	5.425	0	1483955	N.D.	6.272 #
26)	L6 AR-1254-1	0.000	5.285	0	2822161	N.D.	9.515 #
27)	L6 AR-1254-2	6.635	5.485	1996642	7252723	21.928	27.824 #
28)	L6 AR-1254-3	7.015	5.875	1876093	5966235	18.018	13.453 #
29)	L6 AR-1254-4	0.000	6.080	0	5578866	N.D.	17.801 #
30)	L6 AR-1254-5	7.715	6.490	2018172	1863068	20.804	4.451 #
31)	L7 AR-1260-1	7.213	0.000	12623770	0	130.003	N.D. #
32)	L7 AR-1260-2	7.414	6.162	3650665	5016058	31.721	9.603 #
33)	L7 AR-1260-3	7.715	0.000	2018172	0	27.420	N.D. #
36)	L8 AR-1262-1	7.715	6.522	2018172	7842164	18.833	38.938 #
40)	L8 AR-1262-5	0.000	7.854	0	1978179	N.D.	6.021 #
44)	L9 AR-1268-4	0.000	7.854	0	1978179	N.D.	5.530 #
45)	L9 AR-1268-5	0.000	8.124	0	2267858	N.D.	0.852 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR031119\
Data File : PR036058.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Mar 2019 17:41
Operator : SM\SJ
Sample : K1807-39
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampled :
TP-20

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 12 03:51:55 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR030119.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 07 10:54:32 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.408 min
Delta R.T.: -0.004 min
Response: 20556066
Conc: 12.93 ng/ml

Instrument :
ECD_R
ClientSampleId :
TP-20




#1 Tetrachloro-m-xylene

R.T.: 3.481 min
Delta R.T.: -0.005 min
Response: 66113384
Conc: 12.59 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.070 min
Delta R.T.: -0.010 min
Response: 10598929
Conc: 6.63 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.354 min
Delta R.T.: -0.009 min
Response: 38110107
Conc: 5.81 ng/ml



#20 AR-1242-5


R.T.: 0.000 min
Exp R.T.: 6.490 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
TP-20




#20 AR-1242-5

R.T.: 5.285 min
Delta R.T.: -0.038 min
Response: 2822161
Conc: 15.99 ng/ml




#25 AR-1248-5

R.T.: 0.000 min
Exp R.T.: 6.489 min
Response: 0
Conc: N.D.



#25 AR-1248-5


R.T.: 5.425 min
Delta R.T.: 0.061 min
Response: 1483955
Conc: 6.27 ng/ml



#26 AR-1254-1


R.T.: 0.000 min
Exp R.T.: 6.424 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
TP-20




#26 AR-1254-1

R.T.: 5.285 min
Delta R.T.: -0.038 min
Response: 2822161
Conc: 9.51 ng/ml




#27 AR-1254-2

R.T.: 6.635 min
Delta R.T.: -0.006 min
Response: 1996642
Conc: 21.93 ng/ml



#27 AR-1254-2


R.T.: 5.485 min
Delta R.T.: 0.017 min
Response: 7252723
Conc: 27.82 ng/ml



#28 AR-1254-3


R.T.: 7.015 min
Delta R.T.: 0.008 min
Response: 1876093
Conc: 18.02 ng/ml

Instrument :
ECD_R
ClientSampleId :
TP-20




#28 AR-1254-3

R.T.: 5.875 min
Delta R.T.: 0.012 min
Response: 5966235
Conc: 13.45 ng/ml




#29 AR-1254-4

R.T.: 0.000 min
Exp R.T. : 7.291 min
Response: 0
Conc: N.D.



#29 AR-1254-4


R.T.: 6.080 min
Delta R.T.: -0.009 min
Response: 5578866
Conc: 17.80 ng/ml



#30 AR-1254-5


R.T.: 7.715 min
Delta R.T.: 0.007 min
Response: 2018172
Conc: 20.80 ng/ml

Instrument :
ECD_R
ClientSampleId :
TP-20




#30 AR-1254-5

R.T.: 6.490 min
Delta R.T.: -0.008 min
Response: 1863068
Conc: 4.45 ng/ml




#31 AR-1260-1

R.T.: 7.213 min
Delta R.T.: 0.046 min
Response: 12623770
Conc: 130.00 ng/ml



#31 AR-1260-1


R.T.: 0.000 min
Exp R.T.: 5.990 min
Response: 0
Conc: N.D.



#32 AR-1260-2


R.T.: 7.414 min
Delta R.T.: -0.010 min
Response: 3650665
Conc: 31.72 ng/ml

Instrument :
ECD_R
ClientSampleId :
TP-20




#32 AR-1260-2

R.T.: 6.162 min
Delta R.T.: -0.014 min
Response: 5016058
Conc: 9.60 ng/ml




#33 AR-1260-3

R.T.: 7.715 min
Delta R.T.: -0.066 min
Response: 2018172
Conc: 27.42 ng/ml



#33 AR-1260-3


R.T.: 0.000 min
Exp R.T.: 6.326 min
Response: 0
Conc: N.D.



#36 AR-1262-1


R.T.: 7.715 min
Delta R.T.: -0.065 min
Response: 2018172
Conc: 18.83 ng/ml

Instrument :
ECD_R
ClientSampleId :
TP-20




#36 AR-1262-1

R.T.: 6.522 min
Delta R.T.: -0.066 min
Response: 7842164
Conc: 38.94 ng/ml




#40 AR-1262-5

R.T.: 0.000 min
Exp R.T. : 9.356 min
Response: 0
Conc: N.D.



#40 AR-1262-5


R.T.: 7.854 min
Delta R.T.: -0.009 min
Response: 1978179
Conc: 6.02 ng/ml



#44 AR-1268-4


R.T.: 0.000 min
Exp R.T.: 9.357 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
TP-20




#44 AR-1268-4

R.T.: 7.854 min
Delta R.T.: -0.009 min
Response: 1978179
Conc: 5.53 ng/ml



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T.: 9.756 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 8.124 min
Delta R.T.: -0.010 min
Response: 2267858
Conc: 0.85 ng/ml