

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR031220CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Mar 13 09:31:34 2020
 Response Via : Initial Calibration

Calibration Files

400 =PR044931.D 1600 =PR044933.D 800 =PR044932.D
 200 =PR044930.D 100 =PR044929.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.491	1.548	1.301	1.400	1.538	1.456	E6	7.16
2) SA	Decachlorobiphenyl	1.634	1.518	1.557	1.748	1.866	1.664	E6	8.58
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1	6.272	5.898	5.380	6.391	6.985	6.185	E4	9.63
27) L6	AR-1254-2	0.979	0.934	0.845	0.989	1.089	0.967	E5	9.16
28) L6	AR-1254-3	1.037	1.009	0.914	1.035	1.130	1.025	E5	7.54
29) L6	AR-1254-4	8.439	8.032	7.425	8.582	9.364	8.368	E4	8.54
30) L6	AR-1254-5	8.827	8.413	7.807	8.836	9.545	8.686	E4	7.34
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PR044931.D 1600 =PR044933.D 800 =PR044932.D
 200 =PR044930.D 100 =PR044929.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.657	1.673	1.442	1.591	1.760	1.625	E7	7.29
2) SA	Decachlorobiphenyl	1.377	1.378	1.376	1.410	1.439	1.396	E7	2.02
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR031220CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Mar 13 09:31:34 2020
 Response Via : Initial Calibration

Calibration Files

400 =PR044931.D 1600 =PR044933.D 800 =PR044932.D
 200 =PR044930.D 100 =PR044929.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	7.656	7.311	6.640	7.695	8.310	7.523 E5	8.12
27) L6	AR-1254-2	6.836	6.520	5.950	6.897	7.556	6.752 E5	8.67
28) L6	AR-1254-3	1.133	1.116	1.009	1.113	1.234	1.121 E6	7.11
29) L6	AR-1254-4	8.466	8.317	7.605	8.351	9.073	8.362 E5	6.24
30) L6	AR-1254-5	1.062	1.059	0.968	1.011	1.098	1.040 E6	4.84
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range