

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR031220CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Mar 13 10:04:03 2020
 Response Via : Initial Calibration

Calibration Files

400 =PR044936.D 1600 =PR044938.D 800 =PR044937.D
 200 =PR044935.D 100 =PR044934.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.413	1.155	1.452	1.694	1.528	1.448	E6	13.53
2) SA	Decachlorobiphenyl	1.661	1.447	1.571	1.716	1.893	1.658	E6	10.03
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1	4.563	3.668	4.380	5.158	5.232	4.600	E4	13.87
37) L8	AR-1262-2	2.101	1.806	2.066	2.283	2.338	2.119	E5	9.90
38) L8	AR-1262-3	1.461	1.221	1.398	1.596	1.653	1.466	E5	11.67
39) L8	AR-1262-4	1.145	0.950	1.092	1.244	1.280	1.142	E5	11.49
40) L8	AR-1262-5	8.535	7.310	8.077	8.888	9.718	8.506	E4	10.57
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PR044936.D 1600 =PR044938.D 800 =PR044937.D
 200 =PR044935.D 100 =PR044934.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.577	1.260	1.594	1.879	1.769	1.616	E7	14.56
2) SA	Decachlorobiphenyl	1.402	1.313	1.377	1.342	1.469	1.381	E7	4.35
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR031220CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Mar 13 10:04:03 2020
 Response Via : Initial Calibration

Calibration Files

400 =PR044936.D 1600 =PR044938.D 800 =PR044937.D
 200 =PR044935.D 100 =PR044934.D

Compound			400	1600	800	200	100	Avg	%RSD

6) L1	AR-1016-4							0.000	-1.00
7) L1	AR-1016-5							0.000	-1.00
8) L2	AR-1221-1							0.000	-1.00
9) L2	AR-1221-2							0.000	-1.00
10) L2	AR-1221-3							0.000	-1.00
11) L3	AR-1232-1							0.000	-1.00
12) L3	AR-1232-2							0.000	-1.00
13) L3	AR-1232-3							0.000	-1.00
14) L3	AR-1232-4							0.000	-1.00
15) L3	AR-1232-5							0.000	-1.00
16) L4	AR-1242-1							0.000	-1.00
17) L4	AR-1242-2							0.000	-1.00
18) L4	AR-1242-3							0.000	-1.00
19) L4	AR-1242-4							0.000	-1.00
20) L4	AR-1242-5							0.000	-1.00
21) L5	AR-1248-1							0.000	-1.00
22) L5	AR-1248-2							0.000	-1.00
23) L5	AR-1248-3							0.000	-1.00
24) L5	AR-1248-4							0.000	-1.00
25) L5	AR-1248-5							0.000	-1.00
26) L6	AR-1254-1							0.000	-1.00
27) L6	AR-1254-2							0.000	-1.00
28) L6	AR-1254-3							0.000	-1.00
29) L6	AR-1254-4							0.000	-1.00
30) L6	AR-1254-5							0.000	-1.00
31) L7	AR-1260-1							0.000	-1.00
32) L7	AR-1260-2							0.000	-1.00
33) L7	AR-1260-3							0.000	-1.00
34) L7	AR-1260-4							0.000	-1.00
35) L7	AR-1260-5							0.000	-1.00
36) L8	AR-1262-1		5.603	4.769	5.397	5.720	6.196	5.537 E5	9.39
37) L8	AR-1262-2		2.613	2.279	2.611	2.739	2.745	2.597 E6	7.29
38) L8	AR-1262-3		1.066	0.920	1.050	1.109	1.195	1.068 E6	9.34
39) L8	AR-1262-4		1.768	1.553	1.754	1.815	1.896	1.757 E6	7.22
40) L8	AR-1262-5		8.342	7.718	8.258	8.510	8.601	8.286 E5	4.16
41) L9	AR-1268-1							0.000	-1.00
42) L9	AR-1268-2							0.000	-1.00
43) L9	AR-1268-3							0.000	-1.00
44) L9	AR-1268-4							0.000	-1.00
45) L9	AR-1268-5							0.000	-1.00

 (#) = Out of Range