

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR031320\
Data File : PR045037.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Mar 2020 10:27
Operator : AJ\MA
Sample : AR1660CCC400
Misc :
ALS Vial : 3 Sample Multiplier: 1


Instrument :
ECD_R
LabSampleId :
AR1660367

Manual Integrations
APPROVED

Sohil
3/17/2020 8:16:52 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 17 07:04:35 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR031220CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 13 12:07:27 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(31) AR-1260-1 (L7)

R.T.	Response	Conc
7.45	25342946	346.26
7.71	31962717	344.32
8.07	25586457	341.58
8.30	27056882	333.77
8.64	57936903	332.87

(31) AR-1260-1 #2 (L7)

R.T.	Response	Conc
6.51	228654196	312.07
6.70	329476939	305.20
6.85	274872094	316.58
7.32	231841249	302.82
7.56	680966249	299.13

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR031320\
 Data File : PR045037.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Mar 2020 10:27
 Operator : AJ\MA
 Sample : AR1660CCC400
 Misc :
 ALS Vial : 3 Sample Multiplier: 1


Instrument :
 ECD_R
 LabSampleId :
 AR1660367

Manual Integrations
 APPROVED

Sohil
 3/17/2020 8:16:52 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 17 07:04:35 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR031220CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Mar 13 12:07:27 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(31) AR-1260-1 (L7)

R.T.	Response	Conc
7.45	25342946	346.26
7.71	31962717	344.32
8.07	25586457	341.58
8.30	27056882	333.77
8.64	57936903	332.87

(31) AR-1260-1 #2 (L7)

R.T.	Response	Conc
6.51	228654196	312.07
6.70	329476939	305.20
6.85	274872094	316.58
7.32	231841249	302.82
7.56	683691536	300.32

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR031320\
 Data File : PR045037.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Mar 2020 10:27
 Operator : AJ\MA
 Sample : AR1660CCC400
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_R
 LabSampleId :
 AR1660367

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 17 07:04:35 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR031220CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Mar 13 12:07:27 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Manual Integrations
 APPROVED

Sohil
 3/17/2020 8:16:52 AM

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.675	3.938	31212386	325.2E6	21.571	19.777
2) SA Decachlor...	10.548	9.004	49673936	375.5E6	34.819	30.655
Target Compounds						
3) L1 AR-1016-1	5.850	5.026	21955326	146.2E6	386.763	350.386
4) L1 AR-1016-2	5.873	5.044	30931215	202.6E6	389.669	345.377
5) L1 AR-1016-3	5.935	5.222	18764179	112.8E6	398.644	343.251
6) L1 AR-1016-4	6.036	5.262	15352483	86306378	392.408	344.635
7) L1 AR-1016-5	6.328	5.478	14559159	118.4E6	374.014	328.074
31) L7 AR-1260-1	7.452	6.510	25342946	228.7E6	346.263	312.072
32) L7 AR-1260-2	7.707	6.698	31962717	329.5E6	344.318	305.196
33) L7 AR-1260-3	8.067	6.852	25586457	274.9E6	341.578	316.576
34) L7 AR-1260-4	8.304	7.324	27056882	231.8E6	333.770	302.815
35) L7 AR-1260-5	8.638	7.565	57936903	683.7E6	332.869	300.324m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

SJ
 03/27/20