

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR031419\
 Data File : PR036167.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Mar 2019 12:05
 Operator : SM\SJ
 Sample : K1849-03DL 5X
 Misc :
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 H1308-PCDL

Manual Integrations
 APPROVED

Sohil
 3/15/2019 8:30:32 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 14 14:15:12 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR030119.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 07 10:54:32 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.406	3.479	5327686	15794943	3.352	3.007
2) SA Decachlor...	10.065	8.349	11044418	36878221	6.906m	5.618
Target Compounds						
31) L7 AR-1260-1	7.149	5.982	60184626	91954289	619.799	252.506 #
32) L7 AR-1260-2	7.414	6.166	31312062	74714053	272.077	143.030 #
33) L7 AR-1260-3	7.771	6.316	24596535	57161776	334.182	130.595 #
34) L7 AR-1260-4	7.998	6.778	20228430	68386883	214.401	220.878
35) L7 AR-1260-5	8.313	7.021	35251642	239.3E6	209.047	264.998 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR031419\
Data File : PR036167.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Mar 2019 12:05
Operator : SM\SJ
Sample : K1849-03DL 5X
Misc :
ALS Vial : 50 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
H1308-PCDL


Manual Integrations
APPROVED

Sohil
3/15/2019 8:30:32 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 14 14:15:12 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR030119.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 07 10:54:32 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.406 min
Delta R.T.: -0.006 min
Response: 5327686
Conc: 3.35 ng/ml

Instrument :
ECD_R
ClientSampleId :
H1308-PCDL


Manual Integrations
APPROVED

Sohil
3/15/2019 8:30:32 AM




#1 Tetrachloro-m-xylene

R.T.: 3.479 min
Delta R.T.: -0.007 min
Response: 15794943
Conc: 3.01 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.065 min
Delta R.T.: -0.015 min
Response: 11044418
Conc: 6.91 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.349 min
Delta R.T.: -0.014 min
Response: 36878221
Conc: 5.62 ng/ml




#31 AR-1260-1

R.T.: 7.149 min
Delta R.T.: -0.019 min
Response: 60184626
Conc: 619.80 ng/ml

Instrument :
ECD_R
ClientSampleId :
H1308-PCDL


Manual Integrations
APPROVED

Sohil
3/15/2019 8:30:32 AM




#31 AR-1260-1

R.T.: 5.982 min
Delta R.T.: -0.008 min
Response: 91954289
Conc: 252.51 ng/ml




#32 AR-1260-2

R.T.: 7.414 min
Delta R.T.: -0.009 min
Response: 31312062
Conc: 272.08 ng/ml



#32 AR-1260-2

R.T.: 6.166 min
Delta R.T.: -0.010 min
Response: 74714053
Conc: 143.03 ng/ml




#33 AR-1260-3

R.T.: 7.771 min
Delta R.T.: -0.009 min
Response: 24596535
Conc: 334.18 ng/ml

Instrument :
ECD_R
ClientSampleId :
H1308-PCDL


Manual Integrations
APPROVED

Sohil
3/15/2019 8:30:32 AM




#33 AR-1260-3

R.T.: 6.316 min
Delta R.T.: -0.010 min
Response: 57161776
Conc: 130.59 ng/ml




#34 AR-1260-4

R.T.: 7.998 min
Delta R.T.: -0.007 min
Response: 20228430
Conc: 214.40 ng/ml



#34 AR-1260-4

R.T.: 6.778 min
Delta R.T.: -0.012 min
Response: 68386883
Conc: 220.88 ng/ml




#35 AR-1260-5

R.T.: 8.313 min
Delta R.T.: -0.010 min
Response: 35251642
Conc: 209.05 ng/ml

Instrument :
ECD_R
ClientSampleId :
H1308-PCDL

Manual Integrations
APPROVED

Sohil
3/15/2019 8:30:32 AM



#35 AR-1260-5

R.T.: 7.021 min
Delta R.T.: -0.011 min
Response: 239347505
Conc: 265.00 ng/ml