

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR031519\
 Data File : PR036229.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Mar 2019 19:52
 Operator : SM\SJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 16 03:44:58 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR031519.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Mar 16 00:14:35 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.405	3.478	87183926	293.6E6	47.244	46.200
2)	SA Decachlor...	10.064	8.350	72300328	296.4E6	49.506	51.939
Target Compounds							
3)	L1 AR-1016-1	5.569	4.534	30797676	109.7E6	450.991	449.329
4)	L1 AR-1016-2	5.591	4.550	49618921	174.7E6	451.186	454.615
5)	L1 AR-1016-3	5.653	4.722	28556023	87500127	452.630	457.917
6)	L1 AR-1016-4	5.752	4.763	23578192	66316187	454.848	451.846
7)	L1 AR-1016-5	6.043	4.971	24075512	87183035	455.684	447.181
31)	L7 AR-1260-1	7.159	5.980	41266153	165.6E6	449.132	454.244
32)	L7 AR-1260-2	7.415	6.166	48742102	242.9E6	453.325	466.451
33)	L7 AR-1260-3	7.771	6.316	37599914	199.2E6	458.463	469.898
34)	L7 AR-1260-4	7.998	6.780	43130517	171.7E6	463.933	488.322
35)	L7 AR-1260-5	8.312	7.021	84251760	494.8E6	477.552	501.235



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR031519\
Data File : PR036229.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Mar 2019 19:52
Operator : SM\SJ
Sample : AR1660CCC500
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 16 03:44:58 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR031519.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Mar 16 00:14:35 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.405 min
Delta R.T.: 0.000 min
Response: 87183926
Conc: 47.24 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.478 min
Delta R.T.: 0.000 min
Response: 293630282
Conc: 46.20 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.064 min
Delta R.T.: -0.002 min
Response: 72300328
Conc: 49.51 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.350 min
Delta R.T.: 0.000 min
Response: 296395689
Conc: 51.94 ng/ml



#3 AR-1016-1


R.T.: 5.569 min
Delta R.T.: 0.000 min
Response: 30797676
Conc: 450.99 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.534 min
Delta R.T.: 0.000 min
Response: 109688342
Conc: 449.33 ng/ml




#4 AR-1016-2

R.T.: 5.591 min
Delta R.T.: 0.000 min
Response: 49618921
Conc: 451.19 ng/ml



#4 AR-1016-2


R.T.: 4.550 min
Delta R.T.: 0.000 min
Response: 174705708
Conc: 454.62 ng/ml



#5 AR-1016-3


R.T.: 5.653 min
Delta R.T.: 0.000 min
Response: 28556023
Conc: 452.63 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.722 min
Delta R.T.: 0.000 min
Response: 87500127
Conc: 457.92 ng/ml




#6 AR-1016-4

R.T.: 5.752 min
Delta R.T.: 0.000 min
Response: 23578192
Conc: 454.85 ng/ml



#6 AR-1016-4


R.T.: 4.763 min
Delta R.T.: 0.000 min
Response: 66316187
Conc: 451.85 ng/ml



#7 AR-1016-5


R.T.: 6.043 min
Delta R.T.: 0.000 min
Response: 24075512
Conc: 455.68 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 4.971 min
Delta R.T.: 0.000 min
Response: 87183035
Conc: 447.18 ng/ml




#31 AR-1260-1

R.T.: 7.159 min
Delta R.T.: 0.000 min
Response: 41266153
Conc: 449.13 ng/ml



#31 AR-1260-1


R.T.: 5.980 min
Delta R.T.: 0.000 min
Response: 165574439
Conc: 454.24 ng/ml



#32 AR-1260-2


R.T.: 7.415 min
Delta R.T.: 0.000 min
Response: 48742102
Conc: 453.32 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.166 min
Delta R.T.: 0.000 min
Response: 242947805
Conc: 466.45 ng/ml




#33 AR-1260-3

R.T.: 7.771 min
Delta R.T.: 0.000 min
Response: 37599914
Conc: 458.46 ng/ml



#33 AR-1260-3


R.T.: 6.316 min
Delta R.T.: 0.000 min
Response: 199220441
Conc: 469.90 ng/ml



#34 AR-1260-4


R.T.: 7.998 min
Delta R.T.: 0.000 min
Response: 43130517
Conc: 463.93 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 6.780 min
Delta R.T.: 0.000 min
Response: 171669106
Conc: 488.32 ng/ml



#35 AR-1260-5

R.T.: 8.312 min
Delta R.T.: -0.001 min
Response: 84251760
Conc: 477.55 ng/ml



#35 AR-1260-5

R.T.: 7.021 min
Delta R.T.: 0.000 min
Response: 494844518
Conc: 501.24 ng/ml