

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR031919\
 Data File : PR036376.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Mar 2019 17:46
 Operator : SM\SJ
 Sample : K1930-14MS
 Misc :
 ALS Vial : 58 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 TP-6MS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 20 00:47:07 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR031519.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Mar 16 00:14:35 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.406	3.480	34703390	98967672	18.805	15.572
2) SA Decachlor...	10.061	8.347	24257063	69409452	16.609	12.163 #
Target Compounds						
3) L1 AR-1016-1	5.568	4.532	14437812	43721738	211.422	179.102
4) L1 AR-1016-2	5.589	4.549	23309045	64478456	211.950	167.784
5) L1 AR-1016-3	5.651	4.721	13693774	33985620	217.054	177.858
6) L1 AR-1016-4	5.750	4.762	11039420	25752356	212.962	175.464
7) L1 AR-1016-5	6.041	4.970	11034249	33384141	208.848	171.235
31) L7 AR-1260-1	7.157	5.978	20002107	68903680	217.699	189.033
32) L7 AR-1260-2	7.412	6.164	23537800	92490483	218.913	177.578
33) L7 AR-1260-3	7.769	6.313	14653566	80022315	178.674	188.747
34) L7 AR-1260-4	7.995	6.776	17328181	55237211	186.390	157.125
35) L7 AR-1260-5	8.311	7.018	30460944	140.0E6	172.657	141.851



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR031919\
Data File : PR036376.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Mar 2019 17:46
Operator : SM\SJ
Sample : K1930-14MS
Misc :
ALS Vial : 58 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
TP-6MS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 20 00:47:07 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR031519.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Mar 16 00:14:35 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.406 min
Delta R.T.: 0.001 min
Response: 34703390
Conc: 18.81 ng/ml

Instrument :
ECD_R
ClientSampleId :
TP-6MS




#1 Tetrachloro-m-xylene

R.T.: 3.480 min
Delta R.T.: 0.002 min
Response: 98967672
Conc: 15.57 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.061 min
Delta R.T.: -0.006 min
Response: 24257063
Conc: 16.61 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.347 min
Delta R.T.: -0.004 min
Response: 69409452
Conc: 12.16 ng/ml



#3 AR-1016-1


R.T.: 5.568 min
Delta R.T.: -0.002 min
Response: 14437812
Conc: 211.42 ng/ml

Instrument :
ECD_R
ClientSampleId :
TP-6MS




#3 AR-1016-1

R.T.: 4.532 min
Delta R.T.: -0.001 min
Response: 43721738
Conc: 179.10 ng/ml




#4 AR-1016-2

R.T.: 5.589 min
Delta R.T.: -0.002 min
Response: 23309045
Conc: 211.95 ng/ml



#4 AR-1016-2


R.T.: 4.549 min
Delta R.T.: -0.001 min
Response: 64478456
Conc: 167.78 ng/ml



#5 AR-1016-3


R.T.: 5.651 min
Delta R.T.: -0.002 min
Response: 13693774
Conc: 217.05 ng/ml

Instrument :
ECD_R
ClientSampleId :
TP-6MS




#5 AR-1016-3

R.T.: 4.721 min
Delta R.T.: -0.002 min
Response: 33985620
Conc: 177.86 ng/ml




#6 AR-1016-4

R.T.: 5.750 min
Delta R.T.: -0.002 min
Response: 11039420
Conc: 212.96 ng/ml



#6 AR-1016-4


R.T.: 4.762 min
Delta R.T.: -0.002 min
Response: 25752356
Conc: 175.46 ng/ml



#7 AR-1016-5


R.T.: 6.041 min
Delta R.T.: -0.002 min
Response: 11034249
Conc: 208.85 ng/ml

Instrument :
ECD_R
ClientSampleId :
TP-6MS




#7 AR-1016-5

R.T.: 4.970 min
Delta R.T.: -0.001 min
Response: 33384141
Conc: 171.23 ng/ml




#31 AR-1260-1

R.T.: 7.157 min
Delta R.T.: -0.003 min
Response: 20002107
Conc: 217.70 ng/ml



#31 AR-1260-1


R.T.: 5.978 min
Delta R.T.: -0.003 min
Response: 68903680
Conc: 189.03 ng/ml



#32 AR-1260-2


R.T.: 7.412 min
Delta R.T.: -0.003 min
Response: 23537800
Conc: 218.91 ng/ml

Instrument :
ECD_R
ClientSampleId :
TP-6MS




#32 AR-1260-2

R.T.: 6.164 min
Delta R.T.: -0.003 min
Response: 92490483
Conc: 177.58 ng/ml




#33 AR-1260-3

R.T.: 7.769 min
Delta R.T.: -0.002 min
Response: 14653566
Conc: 178.67 ng/ml



#33 AR-1260-3


R.T.: 6.313 min
Delta R.T.: -0.003 min
Response: 80022315
Conc: 188.75 ng/ml



#34 AR-1260-4


R.T.: 7.995 min
Delta R.T.: -0.003 min
Response: 17328181
Conc: 186.39 ng/ml

Instrument :
ECD_R
ClientSampleId :
TP-6MS




#34 AR-1260-4

R.T.: 6.776 min
Delta R.T.: -0.003 min
Response: 55237211
Conc: 157.13 ng/ml



#35 AR-1260-5

R.T.: 8.311 min
Delta R.T.: -0.003 min
Response: 30460944
Conc: 172.66 ng/ml



#35 AR-1260-5

R.T.: 7.018 min
Delta R.T.: -0.003 min
Response: 140042046
Conc: 141.85 ng/ml