

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR032319\
 Data File : PR036627.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Mar 2019 13:46
 Operator : SM\SJ
 Sample : AR1660ICC750
 Misc :
 ALS Vial : 62 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1660ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 25 01:43:57 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR032319.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Mar 25 01:39:01 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.395	3.468	124.3E6	378.1E6	72.699	71.976
2) SA Decachlor...	10.046	8.333	111.3E6	395.6E6	75.046	75.885
Target Compounds						
3) L1 AR-1016-1	5.559	4.522	44492503	139.2E6	725.496	716.866
4) L1 AR-1016-2	5.580	4.539	71397720	230.7E6	711.755	725.473
5) L1 AR-1016-3	5.642	4.711	41114451	109.3E6	709.604	719.348
6) L1 AR-1016-4	5.740	4.752	34172379	84869833	708.826	714.743
7) L1 AR-1016-5	6.032	4.960	34909873	115.1E6	713.579	713.875
31) L7 AR-1260-1	7.148	5.969	62778893	233.2E6	715.300	720.168
32) L7 AR-1260-2	7.403	6.154	77074433	352.7E6	719.430	725.564
33) L7 AR-1260-3	7.760	6.303	59372599	284.8E6	712.032	726.488
34) L7 AR-1260-4	7.986	6.766	69536724	249.6E6	721.516	725.156
35) L7 AR-1260-5	8.300	7.008	140.4E6	729.6E6	740.117	731.826



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR032319\
Data File : PR036627.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Mar 2019 13:46
Operator : SM\SJ
Sample : AR1660ICC750
Misc :
ALS Vial : 62 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 25 01:43:57 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR032319.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Mar 25 01:39:01 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.395 min
Delta R.T.: 0.000 min
Response: 124296524
Conc: 72.70 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.468 min
Delta R.T.: 0.000 min
Response: 378058012
Conc: 71.98 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.046 min
Delta R.T.: 0.000 min
Response: 111313016
Conc: 75.05 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.333 min
Delta R.T.: 0.000 min
Response: 395597522
Conc: 75.89 ng/ml



#3 AR-1016-1


R.T.: 5.559 min
Delta R.T.: 0.000 min
Response: 44492503
Conc: 725.50 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660ICC750




#3 AR-1016-1

R.T.: 4.522 min
Delta R.T.: 0.000 min
Response: 139193078
Conc: 716.87 ng/ml




#4 AR-1016-2

R.T.: 5.580 min
Delta R.T.: 0.000 min
Response: 71397720
Conc: 711.75 ng/ml



#4 AR-1016-2


R.T.: 4.539 min
Delta R.T.: 0.000 min
Response: 230743746
Conc: 725.47 ng/ml



#5 AR-1016-3


R.T.: 5.642 min
Delta R.T.: 0.000 min
Response: 41114451
Conc: 709.60 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660ICC750




#5 AR-1016-3

R.T.: 4.711 min
Delta R.T.: 0.000 min
Response: 109322104
Conc: 719.35 ng/ml




#6 AR-1016-4

R.T.: 5.740 min
Delta R.T.: 0.000 min
Response: 34172379
Conc: 708.83 ng/ml



#6 AR-1016-4


R.T.: 4.752 min
Delta R.T.: 0.000 min
Response: 84869833
Conc: 714.74 ng/ml



#7 AR-1016-5


R.T.: 6.032 min
Delta R.T.: 0.000 min
Response: 34909873
Conc: 713.58 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660ICC750




#7 AR-1016-5

R.T.: 4.960 min
Delta R.T.: 0.000 min
Response: 115131500
Conc: 713.88 ng/ml




#31 AR-1260-1

R.T.: 7.148 min
Delta R.T.: 0.000 min
Response: 62778893
Conc: 715.30 ng/ml



#31 AR-1260-1


R.T.: 5.969 min
Delta R.T.: 0.000 min
Response: 233168851
Conc: 720.17 ng/ml



#32 AR-1260-2


R.T.: 7.403 min
Delta R.T.: 0.000 min
Response: 77074433
Conc: 719.43 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660ICC750




#32 AR-1260-2

R.T.: 6.154 min
Delta R.T.: 0.000 min
Response: 352748558
Conc: 725.56 ng/ml




#33 AR-1260-3

R.T.: 7.760 min
Delta R.T.: 0.000 min
Response: 59372599
Conc: 712.03 ng/ml



#33 AR-1260-3


R.T.: 6.303 min
Delta R.T.: 0.000 min
Response: 284810696
Conc: 726.49 ng/ml



#34 AR-1260-4


R.T.: 7.986 min
Delta R.T.: 0.000 min
Response: 69536724
Conc: 721.52 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660ICC750




#34 AR-1260-4

R.T.: 6.766 min
Delta R.T.: 0.000 min
Response: 249621860
Conc: 725.16 ng/ml



#35 AR-1260-5

R.T.: 8.300 min
Delta R.T.: 0.000 min
Response: 140353894
Conc: 740.12 ng/ml



#35 AR-1260-5

R.T.: 7.008 min
Delta R.T.: 0.000 min
Response: 729639781
Conc: 731.83 ng/ml