

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR032519\
 Data File : PR036658.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Mar 2019 11:58
 Operator : SM\SJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 26 03:22:39 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR032319.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Mar 25 03:18:52 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.397	3.470	39154544	114.9E6	23.245	22.039
2)	SA Decachlor...	10.048	8.334	38726803	118.2E6	25.405	22.756
Target Compounds							
31)	L7 AR-1260-1	7.188	0.000	4680281	0	52.992	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR032519\
Data File : PR036658.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Mar 2019 11:58
Operator : SM\SJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 26 03:22:39 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR032319.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Mar 25 03:18:52 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.397 min
Delta R.T.: 0.000 min
Response: 39154544
Conc: 23.25 ng/ml

Instrument :
ECD_R
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.470 min
Delta R.T.: 0.002 min
Response: 114945433
Conc: 22.04 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.048 min
Delta R.T.: 0.001 min
Response: 38726803
Conc: 25.41 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.334 min
Delta R.T.: 0.000 min
Response: 118240395
Conc: 22.76 ng/ml



#31 AR-1260-1

R.T.: 7.188 min
Delta R.T.: 0.040 min
Response: 4680281
Conc: 52.99 ng/ml

Instrument :
ECD_R
ClientSampleId :
I.BLK



#31 AR-1260-1

R.T.: 0.000 min
Exp R.T. : 5.968 min
Response: 0
Conc: N.D.