

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR040219.M
 Title : GC EXTRACTABLES
 Last Update : Wed Apr 03 03:09:33 2019
 Response Via : Initial Calibration

Calibration Files

750 =PR036764.D 1000 =PR036763.D 500 =PR036765.D
 250 =PR036766.D 50 =PR036767.D

Compound		750	1000	500	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.667	1.345	1.695	1.733	1.526	1.593	E6	9.99
2) SA	Decachlorobiphenyl	1.737	1.767	1.748	1.782	2.334	1.874	E6	13.76
3) L1	AR-1016-1	6.368	5.071	6.488	7.055	7.427	6.482	E4	13.85
4) L1	AR-1016-2	1.042	0.842	1.066	1.122	1.185	1.051	E5	12.29
5) L1	AR-1016-3	5.998	4.886	6.387	6.744	7.558	6.315	E4	15.59
6) L1	AR-1016-4	4.992	4.096	5.297	5.514	6.436	5.267	E4	16.10
7) L1	AR-1016-5	5.146	4.204	5.395	5.881	6.594	5.444	E4	16.28
8) L2	AR-1221-1			1.779			1.779	E4	0.00
9) L2	AR-1221-2			1.350			1.350	E4	0.00
10) L2	AR-1221-3			4.309			4.309	E4	0.00
11) L3	AR-1232-1			2.792			2.792	E4	0.00
12) L3	AR-1232-2			1.562			1.562	E4	0.00
13) L3	AR-1232-3			3.161			3.161	E4	0.00
14) L3	AR-1232-4			1.591			1.591	E4	0.00
15) L3	AR-1232-5			1.247			1.247	E4	0.00
16) L4	AR-1242-1			3.141			3.141	E4	0.00
17) L4	AR-1242-2			5.168			5.168	E4	0.00
18) L4	AR-1242-3			3.190			3.190	E4	0.00
19) L4	AR-1242-4			2.599			2.599	E4	0.00
20) L4	AR-1242-5			3.073			3.073	E4	0.00
21) L5	AR-1248-1			2.578			2.578	E4	0.00
22) L5	AR-1248-2			4.009			4.009	E4	0.00
23) L5	AR-1248-3			4.530			4.530	E4	0.00
24) L5	AR-1248-4			5.373			5.373	E4	0.00
25) L5	AR-1248-5			5.182			5.182	E4	0.00
26) L6	AR-1254-1			5.796			5.796	E4	0.00
27) L6	AR-1254-2			9.338			9.338	E4	0.00
28) L6	AR-1254-3			1.039			1.039	E5	0.00
29) L6	AR-1254-4			7.814			7.814	E4	0.00
30) L6	AR-1254-5			1.022			1.022	E5	0.00
31) L7	AR-1260-1	0.961	0.819	1.013	1.069	1.234	1.019	E5	14.91
32) L7	AR-1260-2	1.167	1.013	1.219	1.262	1.474	1.227	E5	13.62
33) L7	AR-1260-3	0.906	0.800	0.948	0.976	1.169	0.960	E5	14.03
34) L7	AR-1260-4	1.052	0.956	1.092	1.113	1.327	1.108	E5	12.32
35) L7	AR-1260-5	2.117	1.997	2.146	2.126	2.508	2.179	E5	8.86
36) L8	AR-1262-1			1.173			1.173	E5	0.00
37) L8	AR-1262-2			2.156			2.156	E5	0.00
38) L8	AR-1262-3			1.461			1.461	E5	0.00
39) L8	AR-1262-4			1.096			1.096	E5	0.00
40) L8	AR-1262-5			7.470			7.470	E4	0.00
41) L9	AR-1268-1			2.338			2.338	E5	0.00
42) L9	AR-1268-2			2.132			2.132	E5	0.00
43) L9	AR-1268-3			1.789			1.789	E5	0.00
44) L9	AR-1268-4			7.953			7.953	E4	0.00
45) L9	AR-1268-5			5.899			5.899	E5	0.00

Signal #2 Calibration Files

750 =PR036764.D 1000 =PR036763.D 500 =PR036765.D
 250 =PR036766.D 50 =PR036767.D

Compound		750	1000	500	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	6.158	4.936	6.257	6.364	5.574	5.858	E6	10.23
2) SA	Decachlorobiphenyl	7.892	7.797	7.576	7.430	9.217	7.982	E6	8.94
3) L1	AR-1016-1	2.492	1.949	2.496	2.628	2.510	2.415	E5	11.02
4) L1	AR-1016-2	3.954	3.194	4.038	4.197	4.137	3.904	E5	10.44
5) L1	AR-1016-3	1.989	1.593	2.015	2.117	2.039	1.951	E5	10.54

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR040219.M
 Title : GC EXTRACTABLES
 Last Update : Wed Apr 03 03:09:33 2019
 Response Via : Initial Calibration

Calibration Files

750 =PR036764.D 1000 =PR036763.D 500 =PR036765.D
 250 =PR036766.D 50 =PR036767.D

	Compound	750	1000	500	250	50	Avg		%RSD

6)	L1 AR-1016-4	1.520	1.213	1.555	1.646	1.659	1.519	E5	11.89
7)	L1 AR-1016-5	2.065	1.664	2.114	2.230	2.319	2.078	E5	12.12
8)	L2 AR-1221-1			6.141			6.141	E4	0.00
9)	L2 AR-1221-2			4.522			4.522	E4	0.00
10)	L2 AR-1221-3			1.505			1.505	E5	0.00
11)	L3 AR-1232-1			9.455			9.455	E4	0.00
12)	L3 AR-1232-2			1.194			1.194	E5	0.00
13)	L3 AR-1232-3			5.746			5.746	E4	0.00
14)	L3 AR-1232-4			4.947			4.947	E4	0.00
15)	L3 AR-1232-5			5.731			5.731	E4	0.00
16)	L4 AR-1242-1			1.143			1.143	E5	0.00
17)	L4 AR-1242-2			1.902			1.902	E5	0.00
18)	L4 AR-1242-3			9.561			9.561	E4	0.00
19)	L4 AR-1242-4			9.029			9.029	E4	0.00
20)	L4 AR-1242-5			1.381			1.381	E5	0.00
21)	L5 AR-1248-1			9.711			9.711	E4	0.00
22)	L5 AR-1248-2			1.350			1.350	E5	0.00
23)	L5 AR-1248-3			1.392			1.392	E5	0.00
24)	L5 AR-1248-4			1.764			1.764	E5	0.00
25)	L5 AR-1248-5			1.944			1.944	E5	0.00
26)	L6 AR-1254-1			3.289			3.289	E5	0.00
27)	L6 AR-1254-2			2.961			2.961	E5	0.00
28)	L6 AR-1254-3			5.087			5.087	E5	0.00
29)	L6 AR-1254-4			3.832			3.832	E5	0.00
30)	L6 AR-1254-5			5.084			5.084	E5	0.00
31)	L7 AR-1260-1	4.324	3.610	4.453	4.598	4.885	4.374	E5	10.87
32)	L7 AR-1260-2	6.501	5.532	6.629	6.760	7.283	6.541	E5	9.75
33)	L7 AR-1260-3	5.270	4.404	5.403	5.475	5.893	5.289	E5	10.34
34)	L7 AR-1260-4	4.601	4.044	4.735	4.696	5.242	4.663	E5	9.15
35)	L7 AR-1260-5	1.332	1.215	1.355	1.322	1.434	1.331	E6	5.90
36)	L8 AR-1262-1			2.741			2.741	E5	0.00
37)	L8 AR-1262-2			1.371			1.371	E6	0.00
38)	L8 AR-1262-3			5.459			5.459	E5	0.00
39)	L8 AR-1262-4			9.186			9.186	E5	0.00
40)	L8 AR-1262-5			4.017			4.017	E5	0.00
41)	L9 AR-1268-1			1.480			1.480	E6	0.00
42)	L9 AR-1268-2			1.255			1.255	E6	0.00
43)	L9 AR-1268-3			9.726			9.726	E5	0.00
44)	L9 AR-1268-4			4.344			4.344	E5	0.00
45)	L9 AR-1268-5			3.019			3.019	E6	0.00

 (#) = Out of Range