

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR041020\
 Data File : PR045168.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2020 21:48
 Operator : AJ\MA
 Sample : AR1254ICC800
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 11 06:30:37 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR041020CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Apr 11 06:27:37 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.667	3.935	59740109	594.7E6	37.034	36.808
2)	SA Decachlor...	10.536	8.999	108.0E6	863.0E6	76.084	78.976
Target Compounds							
26)	L6 AR-1254-1	6.697	5.826	44035396	477.3E6	698.612	723.544
27)	L6 AR-1254-2	6.914	5.973	68313863	418.2E6	712.837	725.776
28)	L6 AR-1254-3	7.283	6.378	71858864	706.3E6	724.692	741.536
29)	L6 AR-1254-4	7.568	6.605	57914149	519.6E6	721.919	740.791
30)	L6 AR-1254-5	7.987	7.023	59541437	663.7E6	735.921	755.455



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR041020\
Data File : PR045168.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2020 21:48
Operator : AJ\MA
Sample : AR1254ICC800
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 11 06:30:37 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR041020CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Apr 11 06:27:37 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#26 AR-1254-1


R.T.: 6.697 min
Delta R.T.: 0.000 min
Response: 44035396
Conc: 698.61 ng/ml

Instrument :
ECD_R
ClientSampleId :




#26 AR-1254-1

R.T.: 5.826 min
Delta R.T.: 0.000 min
Response: 477339002
Conc: 723.54 ng/ml




#27 AR-1254-2

R.T.: 6.914 min
Delta R.T.: 0.000 min
Response: 68313863
Conc: 712.84 ng/ml



#27 AR-1254-2


R.T.: 5.973 min
Delta R.T.: 0.000 min
Response: 418217063
Conc: 725.78 ng/ml



#28 AR-1254-3


R.T.: 7.283 min
Delta R.T.: 0.000 min
Response: 71858864
Conc: 724.69 ng/ml

Instrument :
ECD_R
ClientSampleId :




#28 AR-1254-3

R.T.: 6.378 min
Delta R.T.: 0.000 min
Response: 706343166
Conc: 741.54 ng/ml




#29 AR-1254-4

R.T.: 7.568 min
Delta R.T.: 0.000 min
Response: 57914149
Conc: 721.92 ng/ml



#29 AR-1254-4


R.T.: 6.605 min
Delta R.T.: 0.000 min
Response: 519639535
Conc: 740.79 ng/ml



#30 AR-1254-5

R.T.: 7.987 min
Delta R.T.: 0.000 min
Response: 59541437
Conc: 735.92 ng/ml

Instrument :
ECD_R
ClientSampleId :



#30 AR-1254-5

R.T.: 7.023 min
Delta R.T.: 0.000 min
Response: 663685577
Conc: 755.46 ng/ml