

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR041819\
 Data File : PR037218.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Apr 2019 17:35
 Operator : SM\SJ
 Sample : K1560-07
 Misc : AR1254 LOD 40PPB
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 LOD-MDL-WATER-01-QT1-2019

Manual Integrations
 APPROVED

Sohil
 4/23/2019 12:03:31 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 19 07:59:23 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR040219.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 03 03:09:33 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.384	3.457	22860899	93526856	14.349	15.966
2)	SA Decachlor...	10.017	8.306	29483079	123.2E6	15.737	15.435
Target Compounds							
26)	L6 AR-1254-1	6.388	5.283	2594820	14154963	44.769m	43.039
27)	L6 AR-1254-2	6.605	5.427	3523593	12155163	37.733	41.057
28)	L6 AR-1254-3	6.971	5.820	3576382	17125517	34.432	33.664
29)	L6 AR-1254-4	7.255	6.046	2829891	14110176	36.215	36.820
30)	L6 AR-1254-5	7.670	6.452	3085202	14412359	30.199m	28.347m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR041819\
Data File : PR037218.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Apr 2019 17:35
Operator : SM\SJ
Sample : K1560-07
Misc : AR1254 LOD 40PPB
ALS Vial : 45 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT1-2019

Manual Integrations
APPROVED

Sohil
4/23/2019 12:03:31 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 19 07:59:23 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR040219.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 03 03:09:33 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

