

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR042219.M
 Title : GC EXTRACTABLES
 Last Update : Mon Apr 22 17:27:39 2019
 Response Via : Initial Calibration

Calibration Files

750 =PR037265.D 1000 =PR037264.D 500 =PR037266.D
 250 =PR037267.D 50 =PR037268.D

Compound		750	1000	500	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.292	1.267	1.231	1.256	1.210	1.251	E6	2.54
2) SA	Decachlorobiphenyl	1.222	1.270	1.276	1.366	1.479	1.323	E6	7.70
3) L1	AR-1016-1	3.962	4.189	3.816	4.396	5.182	4.309	E4	12.43
4) L1	AR-1016-2	6.597	6.889	6.385	7.217	7.536	6.925	E4	6.69
5) L1	AR-1016-3	3.733	3.948	3.661	4.211	4.611	4.033	E4	9.61
6) L1	AR-1016-4	3.032	3.230	2.956	3.361	3.841	3.284	E4	10.66
7) L1	AR-1016-5	3.108	3.353	3.057	3.602	4.025	3.429	E4	11.59
8) L2	AR-1221-1			1.250			1.250	E4	0.00
9) L2	AR-1221-2			9.044			9.044	E3	0.00
10) L2	AR-1221-3			3.065			3.065	E4	0.00
11) L3	AR-1232-1			2.925			2.925	E4	0.00
12) L3	AR-1232-2			1.690			1.690	E4	0.00
13) L3	AR-1232-3			3.380			3.380	E4	0.00
14) L3	AR-1232-4			1.570			1.570	E4	0.00
15) L3	AR-1232-5			1.170			1.170	E4	0.00
16) L4	AR-1242-1			3.421			3.421	E4	0.00
17) L4	AR-1242-2			5.547			5.547	E4	0.00
18) L4	AR-1242-3			3.291			3.291	E4	0.00
19) L4	AR-1242-4			2.638			2.638	E4	0.00
20) L4	AR-1242-5			3.218			3.218	E4	0.00
21) L5	AR-1248-1			2.291			2.291	E4	0.00
22) L5	AR-1248-2			3.501			3.501	E4	0.00
23) L5	AR-1248-3			4.091			4.091	E4	0.00
24) L5	AR-1248-4			4.895			4.895	E4	0.00
25) L5	AR-1248-5			4.724			4.724	E4	0.00
26) L6	AR-1254-1			5.463			5.463	E4	0.00
27) L6	AR-1254-2			8.570			8.570	E4	0.00
28) L6	AR-1254-3			9.096			9.096	E4	0.00
29) L6	AR-1254-4			6.701			6.701	E4	0.00
30) L6	AR-1254-5			7.188			7.188	E4	0.00
31) L7	AR-1260-1	5.667	6.549	5.712	7.203	7.862	6.599	E4	14.41
32) L7	AR-1260-2	6.804	8.008	6.941	8.715	9.384	7.970	E4	13.99
33) L7	AR-1260-3	5.323	6.305	5.487	6.893	7.279	6.258	E4	13.65
34) L7	AR-1260-4	6.242	7.430	6.447	8.128	8.955	7.440	E4	15.31
35) L7	AR-1260-5	1.272	1.517	1.298	1.552	1.592	1.446	E5	10.36
36) L8	AR-1262-1			1.045			1.045	E5	0.00
37) L8	AR-1262-2			1.944			1.944	E5	0.00
38) L8	AR-1262-3			1.315			1.315	E5	0.00
39) L8	AR-1262-4			9.823			9.823	E4	0.00
40) L8	AR-1262-5			6.904			6.904	E4	0.00
41) L9	AR-1268-1			1.791			1.791	E5	0.00
42) L9	AR-1268-2			1.645			1.645	E5	0.00
43) L9	AR-1268-3			1.384			1.384	E5	0.00
44) L9	AR-1268-4			6.239			6.239	E4	0.00
45) L9	AR-1268-5			4.559			4.559	E5	0.00

Signal #2 Calibration Files

750 =PR037265.D 1000 =PR037264.D 500 =PR037266.D
 250 =PR037267.D 50 =PR037268.D

Compound		750	1000	500	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	4.846	4.748	4.679	4.734	4.970	4.796	E6	2.39
2) SA	Decachlorobiphenyl	5.636	5.794	5.806	6.031	5.855	5.824	E6	2.43
3) L1	AR-1016-1	1.705	1.781	1.679	1.873	2.067	1.821	E5	8.62
4) L1	AR-1016-2	2.628	2.797	2.552	2.949	3.179	2.821	E5	8.94
5) L1	AR-1016-3	1.313	1.410	1.290	1.482	1.694	1.438	E5	11.30

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR042219.M
 Title : GC EXTRACTABLES
 Last Update : Mon Apr 22 17:27:39 2019
 Response Via : Initial Calibration

Calibration Files

750 =PR037265.D 1000 =PR037264.D 500 =PR037266.D
 250 =PR037267.D 50 =PR037268.D

	Compound	750	1000	500	250	50	Avg		%RSD
6)	L1 AR-1016-4	0.998	1.070	0.990	1.169	1.305	1.106	E5	11.93
7)	L1 AR-1016-5	1.347	1.446	1.329	1.601	1.878	1.520	E5	14.95
8)	L2 AR-1221-1			5.092			5.092	E4	0.00
9)	L2 AR-1221-2			3.714			3.714	E4	0.00
10)	L2 AR-1221-3			1.224			1.224	E5	0.00
11)	L3 AR-1232-1			1.170			1.170	E5	0.00
12)	L3 AR-1232-2			1.382			1.382	E5	0.00
13)	L3 AR-1232-3			7.020			7.020	E4	0.00
14)	L3 AR-1232-4			5.967			5.967	E4	0.00
15)	L3 AR-1232-5			6.824			6.824	E4	0.00
16)	L4 AR-1242-1			1.465			1.465	E5	0.00
17)	L4 AR-1242-2			2.282			2.282	E5	0.00
18)	L4 AR-1242-3			1.156			1.156	E5	0.00
19)	L4 AR-1242-4			1.099			1.099	E5	0.00
20)	L4 AR-1242-5			1.657			1.657	E5	0.00
21)	L5 AR-1248-1			1.007			1.007	E5	0.00
22)	L5 AR-1248-2			1.365			1.365	E5	0.00
23)	L5 AR-1248-3			1.400			1.400	E5	0.00
24)	L5 AR-1248-4			1.780			1.780	E5	0.00
25)	L5 AR-1248-5			1.956			1.956	E5	0.00
26)	L6 AR-1254-1			3.445			3.445	E5	0.00
27)	L6 AR-1254-2			2.995			2.995	E5	0.00
28)	L6 AR-1254-3			4.736			4.736	E5	0.00
29)	L6 AR-1254-4			3.506			3.506	E5	0.00
30)	L6 AR-1254-5			4.428			4.428	E5	0.00
31)	L7 AR-1260-1	2.538	2.906	2.586	3.220	3.363	2.923	E5	12.62
32)	L7 AR-1260-2	3.847	4.426	3.939	4.998	5.819	4.606	E5	17.77
33)	L7 AR-1260-3	3.081	3.594	3.144	3.926	4.080	3.565	E5	12.61
34)	L7 AR-1260-4	2.658	3.123	2.758	3.394	3.305	3.048	E5	10.73
35)	L7 AR-1260-5	7.804	9.062	7.969	9.433	9.095	8.672	E5	8.47
36)	L8 AR-1262-1			2.506			2.506	E5	0.00
37)	L8 AR-1262-2			1.256			1.256	E6	0.00
38)	L8 AR-1262-3			5.483			5.483	E5	0.00
39)	L8 AR-1262-4			8.976			8.976	E5	0.00
40)	L8 AR-1262-5			3.826			3.826	E5	0.00
41)	L9 AR-1268-1			1.231			1.231	E6	0.00
42)	L9 AR-1268-2			1.059			1.059	E6	0.00
43)	L9 AR-1268-3			8.862			8.862	E5	0.00
44)	L9 AR-1268-4			3.628			3.628	E5	0.00
45)	L9 AR-1268-5			2.544			2.544	E6	0.00

(#) = Out of Range