

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR050219\
 Data File : PR037537.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 May 2019 09:33
 Operator : SM\SJ
 Sample : K2199-17
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 OR-02-042919-A

Manual Integrations
 APPROVED

Ankita
 5/3/2019 3:26:47 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 02 09:49:22 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR042219.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Apr 22 17:27:39 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

1) SA Tetrachlo...	4.377	3.407	23645373	90584821	18.901m	18.889
2) SA Decachlor...	10.010	8.249	18075751	63183368	13.667	10.848

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR050219\
Data File : PR037537.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 May 2019 09:33
Operator : SM\SJ
Sample : K2199-17
Misc :
ALS Vial : 4 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
OR-02-042919-A


Manual Integrations
APPROVED

Ankita
5/3/2019 3:26:47 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 02 09:49:22 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR042219.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Apr 22 17:27:39 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.377 min
Delta R.T.: 0.015 min
Response: 23645373
Conc: 18.90 ng/ml m

Instrument :
ECD_R
ClientSampleId :
OR-02-042919-A


Manual Integrations
APPROVED

Ankita
5/3/2019 3:26:47 PM




#1 Tetrachloro-m-xylene

R.T.: 3.407 min
Delta R.T.: 0.000 min
Response: 90584821
Conc: 18.89 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.010 min
Delta R.T.: 0.013 min
Response: 18075751
Conc: 13.67 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.249 min
Delta R.T.: 0.002 min
Response: 63183368
Conc: 10.85 ng/ml