

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR050319\
 Data File : PR037586.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 May 2019 12:05
 Operator : SM\SJ
 Sample : PB119287BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 PB119287BL

Manual Integrations
 APPROVED

Ankita
 5/6/2019 9:54:50 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 03 12:55:33 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR042219.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Apr 22 17:27:39 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.360	3.406	22842226	91265762	18.259m	19.031
2) SA Decachlor...	9.984	8.241	32444667	134.7E6	24.531m	23.134

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR050319\
Data File : PR037586.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 May 2019 12:05
Operator : SM\SJ
Sample : PB119287BL
Misc :
ALS Vial : 10 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
PB119287BL


Manual Integrations
APPROVED

Ankita
5/6/2019 9:54:50 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 03 12:55:33 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR042219.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Apr 22 17:27:39 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.360 min
Delta R.T.: -0.002 min
Response: 22842226
Conc: 18.26 ng/ml m

Instrument :
ECD_R
ClientSampleId :
PB119287BL


Manual Integrations
APPROVED

Ankita
5/6/2019 9:54:50 AM




#1 Tetrachloro-m-xylene

R.T.: 3.406 min
Delta R.T.: 0.000 min
Response: 91265762
Conc: 19.03 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.984 min
Delta R.T.: -0.013 min
Response: 32444667
Conc: 24.53 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.241 min
Delta R.T.: -0.006 min
Response: 134736099
Conc: 23.13 ng/ml