

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR051019\
 Data File : PR037727.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 May 2019 19:03
 Operator : SM\MA
 Sample : K2642-03
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 OK-03-050919

Manual Integrations
 APPROVED

Ankita
 5/13/2019 2:07:57 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 10 23:07:19 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR050819.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed May 08 18:26:47 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.785	4.629	34953696	104.1E6	20.497m	20.987
2) SA Decachlor...	8.783	10.389	22762509	64760492	9.640	13.652 #

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR051019\
Data File : PR037727.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 May 2019 19:03
Operator : SM\MA
Sample : K2642-03
Misc :
ALS Vial : 23 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
OK-03-050919


Manual Integrations
APPROVED

Ankita
5/13/2019 2:07:57 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 10 23:07:19 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR050819.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed May 08 18:26:47 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.785 min
Delta R.T.: -0.001 min
Response: 34953696
Conc: 20.50 ng/ml m

Instrument :
ECD_R
ClientSampleId :
OK-03-050919


Manual Integrations
APPROVED

Ankita
5/13/2019 2:07:57 PM




#1 Tetrachloro-m-xylene

R.T.: 4.629 min
Delta R.T.: -0.003 min
Response: 104099947
Conc: 20.99 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.783 min
Delta R.T.: -0.013 min
Response: 22762509
Conc: 9.64 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.389 min
Delta R.T.: -0.017 min
Response: 64760492
Conc: 13.65 ng/ml