

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR051119.M
 Title : GC EXTRACTABLES
 Last Update : Sun May 12 00:51:44 2019
 Response Via : Initial Calibration

Calibration Files

750 =PR037756.D 1000 =PR037755.D 500 =PR037757.D
 250 =PR037758.D 50 =PR037759.D

Compound		750	1000	500	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.941	2.020	1.931	1.625	1.493	1.802	E6	12.72
2) SA	Decachlorobiphenyl	1.857	1.542	1.868	2.087	2.185	1.908	E6	13.03
3) L1	AR-1016-1	7.966	8.168	8.226	7.408	7.217	7.797	E4	5.87
4) L1	AR-1016-2	1.111	1.139	1.137	1.031	0.971	1.078	E5	6.89
5) L1	AR-1016-3	5.880	5.955	5.995	5.495	5.036	5.672	E4	7.18
6) L1	AR-1016-4	4.579	4.615	4.735	4.420	4.503	4.570	E4	2.60
7) L1	AR-1016-5	6.237	6.264	6.467	6.044	6.000	6.202	E4	3.03
8) L2	AR-1221-1			2.088			2.088	E4	0.00
9) L2	AR-1221-2			1.576			1.576	E4	0.00
10) L2	AR-1221-3			5.043			5.043	E4	0.00
11) L3	AR-1232-1			5.509			5.509	E4	0.00
12) L3	AR-1232-2			6.802			6.802	E4	0.00
13) L3	AR-1232-3			3.534			3.534	E4	0.00
14) L3	AR-1232-4			3.085			3.085	E4	0.00
15) L3	AR-1232-5			3.512			3.512	E4	0.00
16) L4	AR-1242-1			6.068			6.068	E4	0.00
17) L4	AR-1242-2			8.331			8.331	E4	0.00
18) L4	AR-1242-3			4.447			4.447	E4	0.00
19) L4	AR-1242-4			4.286			4.286	E4	0.00
20) L4	AR-1242-5			6.223			6.223	E4	0.00
21) L5	AR-1248-1			4.769			4.769	E4	0.00
22) L5	AR-1248-2			6.170			6.170	E4	0.00
23) L5	AR-1248-3			6.334			6.334	E4	0.00
24) L5	AR-1248-4			8.058			8.058	E4	0.00
25) L5	AR-1248-5			8.497			8.497	E4	0.00
26) L6	AR-1254-1			1.480			1.480	E5	0.00
27) L6	AR-1254-2			1.288			1.288	E5	0.00
28) L6	AR-1254-3			2.207			2.207	E5	0.00
29) L6	AR-1254-4			1.514			1.514	E5	0.00
30) L6	AR-1254-5			1.959			1.959	E5	0.00
31) L7	AR-1260-1	1.228	1.144	1.271	1.194	1.163	1.200	E5	4.26
32) L7	AR-1260-2	1.569	1.420	1.619	1.543	1.687	1.568	E5	6.31
33) L7	AR-1260-3	1.364	1.233	1.397	1.312	1.304	1.322	E5	4.73
34) L7	AR-1260-4	1.098	0.949	1.123	1.078	1.080	1.065	E5	6.37
35) L7	AR-1260-5	2.956	2.484	2.984	2.854	2.769	2.809	E5	7.15
36) L8	AR-1262-1			6.685			6.685	E4	0.00
37) L8	AR-1262-2			3.266			3.266	E5	0.00
38) L8	AR-1262-3			1.253			1.253	E5	0.00
39) L8	AR-1262-4			2.325			2.325	E5	0.00
40) L8	AR-1262-5			1.073			1.073	E5	0.00
41) L9	AR-1268-1			4.357			4.357	E5	0.00
42) L9	AR-1268-2			3.933			3.933	E5	0.00
43) L9	AR-1268-3			3.217			3.217	E5	0.00
44) L9	AR-1268-4			1.283			1.283	E5	0.00
45) L9	AR-1268-5			9.320			9.320	E5	0.00

Signal #2 Calibration Files

750 =PR037756.D 1000 =PR037755.D 500 =PR037757.D
 250 =PR037758.D 50 =PR037759.D

Compound		750	1000	500	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	5.409	5.619	5.422	4.624	4.140	5.043	E6	12.55
2) SA	Decachlorobiphenyl	4.489	3.751	4.513	4.962	5.025	4.548	E6	11.20
3) L1	AR-1016-1	1.940	2.010	1.997	1.774	1.684	1.881	E5	7.69
4) L1	AR-1016-2	2.804	2.885	2.875	2.573	2.406	2.709	E5	7.78
5) L1	AR-1016-3	1.675	1.699	1.727	1.560	1.480	1.628	E5	6.41

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR051119.M
 Title : GC EXTRACTABLES
 Last Update : Sun May 12 00:51:44 2019
 Response Via : Initial Calibration

Calibration Files

750 =PR037756.D 1000 =PR037755.D 500 =PR037757.D
 250 =PR037758.D 50 =PR037759.D

Compound			750	1000	500	250	50	Avg		%RSD

6)	L1	AR-1016-4	1.414	1.442	1.449	1.300	1.278	1.377	E5	5.93
7)	L1	AR-1016-5	1.393	1.401	1.437	1.294	1.267	1.358	E5	5.41
8)	L2	AR-1221-1			5.289			5.289	E4	0.00
9)	L2	AR-1221-2			3.879			3.879	E4	0.00
10)	L2	AR-1221-3			1.290			1.290	E5	0.00
11)	L3	AR-1232-1			1.444			1.444	E5	0.00
12)	L3	AR-1232-2			7.855			7.855	E4	0.00
13)	L3	AR-1232-3			1.697			1.697	E5	0.00
14)	L3	AR-1232-4			8.522			8.522	E4	0.00
15)	L3	AR-1232-5			6.028			6.028	E4	0.00
16)	L4	AR-1242-1			1.454			1.454	E5	0.00
17)	L4	AR-1242-2			2.102			2.102	E5	0.00
18)	L4	AR-1242-3			1.275			1.275	E5	0.00
19)	L4	AR-1242-4			1.069			1.069	E5	0.00
20)	L4	AR-1242-5			1.237			1.237	E5	0.00
21)	L5	AR-1248-1			1.138			1.138	E5	0.00
22)	L5	AR-1248-2			1.550			1.550	E5	0.00
23)	L5	AR-1248-3			1.823			1.823	E5	0.00
24)	L5	AR-1248-4			2.186			2.186	E5	0.00
25)	L5	AR-1248-5			2.082			2.082	E5	0.00
26)	L6	AR-1254-1			2.408			2.408	E5	0.00
27)	L6	AR-1254-2			3.818			3.818	E5	0.00
28)	L6	AR-1254-3			4.226			4.226	E5	0.00
29)	L6	AR-1254-4			3.090			3.090	E5	0.00
30)	L6	AR-1254-5			3.341			3.341	E5	0.00
31)	L7	AR-1260-1	2.518	2.384	2.566	2.371	2.535	2.475	E5	3.67
32)	L7	AR-1260-2	3.061	2.837	3.117	2.889	2.802	2.941	E5	4.76
33)	L7	AR-1260-3	2.270	2.033	2.314	2.166	2.134	2.183	E5	5.12
34)	L7	AR-1260-4	2.653	2.321	2.690	2.552	2.561	2.555	E5	5.63
35)	L7	AR-1260-5	5.531	4.754	5.533	5.229	4.928	5.195	E5	6.76
36)	L8	AR-1262-1			3.160			3.160	E5	0.00
37)	L8	AR-1262-2			6.231			6.231	E5	0.00
38)	L8	AR-1262-3			4.191			4.191	E5	0.00
39)	L8	AR-1262-4			3.191			3.191	E5	0.00
40)	L8	AR-1262-5			2.490			2.490	E5	0.00
41)	L9	AR-1268-1			1.002			1.002	E6	0.00
42)	L9	AR-1268-2			9.040			9.040	E5	0.00
43)	L9	AR-1268-3			7.810			7.810	E5	0.00
44)	L9	AR-1268-4			3.400			3.400	E5	0.00
45)	L9	AR-1268-5			2.552			2.552	E6	0.00

 (#) = Out of Range