

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR051719\
 Data File : PR038040.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 May 2019 22:06
 Operator : SM\MA
 Sample : K2894-17
 Misc :
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 SB-02-9-10

Manual Integrations
 APPROVED

Ankita
 5/20/2019 4:05:59 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 18 00:10:59 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR051119.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sun May 12 00:51:44 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.781	4.628	21257345	68447249	11.796	13.573m
2) SA Decachlor...	8.766	10.385	25013749	67564187	13.111	14.855

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR051719\
Data File : PR038040.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 May 2019 22:06
Operator : SM\MA
Sample : K2894-17
Misc :
ALS Vial : 47 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
SB-02-9-10


Manual Integrations
APPROVED

Ankita
5/20/2019 4:05:59 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 18 00:10:59 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR051119.M
Quant Title : GC EXTRACTABLES
QLast Update : Sun May 12 00:51:44 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.781 min
Delta R.T.: 0.000 min
Response: 21257345
Conc: 11.80 ng/ml

Instrument :
ECD_R
ClientSampleId :
SB-02-9-10


Manual Integrations
APPROVED

Ankita
5/20/2019 4:05:59 PM




#1 Tetrachloro-m-xylene

R.T.: 4.628 min
Delta R.T.: 0.000 min
Response: 68447249
Conc: 13.57 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.766 min
Delta R.T.: -0.009 min
Response: 25013749
Conc: 13.11 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.385 min
Delta R.T.: -0.003 min
Response: 67564187
Conc: 14.86 ng/ml