

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR052319\
 Data File : PR038228.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 May 2019 16:34
 Operator : SM\MA
 Sample : K2995-03RE
 Misc :
 ALS Vial : 146 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 POS-EFFL5-UFRE

Manual Integrations
 APPROVED

Ankita
 5/24/2019 3:12:20 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 23 16:56:20 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR052319.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 23 13:10:55 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.784	4.626	38285733	107.8E6	22.831m	21.256m
2) SA Decachlor...	8.758	10.380	10576764	31574610	5.452	7.158 #

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR052319\
Data File : PR038228.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 May 2019 16:34
Operator : SM\MA
Sample : K2995-03RE
Misc :
ALS Vial : 146 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
POS-EFFL5-UFRE


Manual Integrations
APPROVED

Ankita
5/24/2019 3:12:20 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 23 16:56:20 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR052319.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 23 13:10:55 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.784 min
Delta R.T.: 0.006 min
Response: 38285733
Conc: 22.83 ng/ml m

Instrument :
ECD_R
ClientSampleId :
POS-EFFL5-UFRE


Manual Integrations
APPROVED

Ankita
5/24/2019 3:12:20 PM




#1 Tetrachloro-m-xylene

R.T.: 4.626 min
Delta R.T.: 0.000 min
Response: 107840554
Conc: 21.26 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.758 min
Delta R.T.: -0.001 min
Response: 10576764
Conc: 5.45 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.380 min
Delta R.T.: -0.001 min
Response: 31574610
Conc: 7.16 ng/ml