

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR053119\
 Data File : PR038395.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 May 2019 09:18
 Operator : SM\MA
 Sample : K3084-03
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 TEST-PIT-7

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 01 04:53:33 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR053019.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 30 14:26:18 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.773	4.621	40490589	122.9E6	25.501	24.106
2)	SA Decachlor...	8.752	10.371	37588638	97822451	23.915	23.267
Target Compounds							
35)	L7 AR-1260-5	0.000	8.578	0	14353115	N.D.	32.350 #
37)	L8 AR-1262-2	0.000	8.578	0	14353115	N.D.	22.954 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR053119\
Data File : PR038395.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 May 2019 09:18
Operator : SM\MA
Sample : K3084-03
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
TEST-PIT-7

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 01 04:53:33 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR053019.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 30 14:26:18 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.773 min
Delta R.T.: 0.000 min
Response: 40490589
Conc: 25.50 ng/ml

Instrument :
ECD_R
ClientSampleId :
TEST-PIT-7




#1 Tetrachloro-m-xylene

R.T.: 4.621 min
Delta R.T.: -0.001 min
Response: 122871091
Conc: 24.11 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.752 min
Delta R.T.: -0.001 min
Response: 37588638
Conc: 23.91 ng/ml



#2 Decachlorobiphenyl


R.T.: 10.371 min
Delta R.T.: -0.001 min
Response: 97822451
Conc: 23.27 ng/ml



#35 AR-1260-5


R.T.: 0.000 min
Exp R.T.: 7.358 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
TEST-PIT-7




#35 AR-1260-5

R.T.: 8.578 min
Delta R.T.: 0.053 min
Response: 14353115
Conc: 32.35 ng/ml



#37 AR-1262-2

R.T.: 0.000 min
Exp R.T.: 7.358 min
Response: 0
Conc: N.D.



#37 AR-1262-2

R.T.: 8.578 min
Delta R.T.: 0.053 min
Response: 14353115
Conc: 22.95 ng/ml