

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR060519.M
 Title : GC EXTRACTABLES
 Last Update : Thu Jun 06 01:26:25 2019
 Response Via : Initial Calibration

Calibration Files

750 =PR038492.D 1000 =PR038491.D 500 =PR038493.D
 250 =PR038494.D 50 =PR038495.D

Compound		750	1000	500	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	2.265	2.121	2.088	1.989	2.192	2.131	E6	4.91
2) SA	Decachlorobiphenyl	2.330	2.558	2.371	2.304	2.931	2.499	E6	10.47
3) L1	AR-1016-1	0.807	0.786	0.755	0.747	1.011	0.821	E5	13.23
4) L1	AR-1016-2	1.332	1.331	1.226	1.221	1.562	1.334	E5	10.35
5) L1	AR-1016-3	6.426	6.477	5.964	6.074	7.416	6.472	E4	8.84
6) L1	AR-1016-4	5.039	5.004	4.743	4.832	6.315	5.187	E4	12.39
7) L1	AR-1016-5	6.577	6.587	6.143	6.297	8.556	6.832	E4	14.38
8) L2	AR-1221-1			4.023			4.023	E4	0.00
9) L2	AR-1221-2			3.080			3.080	E4	0.00
10) L2	AR-1221-3			9.854			9.854	E4	0.00
11) L3	AR-1232-1			7.103			7.103	E4	0.00
12) L3	AR-1232-2			8.692			8.692	E4	0.00
13) L3	AR-1232-3			4.146			4.146	E4	0.00
14) L3	AR-1232-4			3.601			3.601	E4	0.00
15) L3	AR-1232-5			4.071			4.071	E4	0.00
16) L4	AR-1242-1			1.062			1.062	E5	0.00
17) L4	AR-1242-2			1.766			1.766	E5	0.00
18) L4	AR-1242-3			8.481			8.481	E4	0.00
19) L4	AR-1242-4			8.079			8.079	E4	0.00
20) L4	AR-1242-5			1.171			1.171	E5	0.00
21) L5	AR-1248-1			8.777			8.777	E4	0.00
22) L5	AR-1248-2			1.172			1.172	E5	0.00
23) L5	AR-1248-3			1.209			1.209	E5	0.00
24) L5	AR-1248-4			1.508			1.508	E5	0.00
25) L5	AR-1248-5			1.620			1.620	E5	0.00
26) L6	AR-1254-1			2.210			2.210	E5	0.00
27) L6	AR-1254-2			1.896			1.896	E5	0.00
28) L6	AR-1254-3			3.149			3.149	E5	0.00
29) L6	AR-1254-4			2.151			2.151	E5	0.00
30) L6	AR-1254-5			2.837			2.837	E5	0.00
31) L7	AR-1260-1	1.215	1.388	1.155	1.145	1.706	1.322	E5	17.83
32) L7	AR-1260-2	1.549	1.842	1.495	1.460	2.181	1.706	E5	17.92
33) L7	AR-1260-3	1.388	1.642	1.334	1.296	1.888	1.510	E5	16.63
34) L7	AR-1260-4	1.164	1.410	1.143	1.128	1.619	1.293	E5	16.69
35) L7	AR-1260-5	3.153	3.916	3.115	2.885	3.971	3.408	E5	14.67
36) L8	AR-1262-1			1.197			1.197	E5	0.00
37) L8	AR-1262-2			6.451			6.451	E5	0.00
38) L8	AR-1262-3			2.727			2.727	E5	0.00
39) L8	AR-1262-4			4.894			4.894	E5	0.00
40) L8	AR-1262-5			2.203			2.203	E5	0.00
41) L9	AR-1268-1			5.179			5.179	E5	0.00
42) L9	AR-1268-2			4.680			4.680	E5	0.00
43) L9	AR-1268-3			3.870			3.870	E5	0.00
44) L9	AR-1268-4			1.617			1.617	E5	0.00
45) L9	AR-1268-5			1.218			1.218	E6	0.00

Signal #2 Calibration Files

750 =PR038492.D 1000 =PR038491.D 500 =PR038493.D
 250 =PR038494.D 50 =PR038495.D

Compound		750	1000	500	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	7.422	6.733	6.867	6.763	7.587	7.074	E6	5.65
2) SA	Decachlorobiphenyl	5.611	6.014	5.803	5.787	7.554	6.154	E6	12.93
3) L1	AR-1016-1	2.661	2.575	2.461	2.484	3.218	2.680	E5	11.61
4) L1	AR-1016-2	3.869	3.750	3.517	3.532	4.502	3.834	E5	10.48
5) L1	AR-1016-3	2.248	2.190	2.079	2.118	2.764	2.280	E5	12.21

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR060519.M
 Title : GC EXTRACTABLES
 Last Update : Thu Jun 06 01:26:25 2019
 Response Via : Initial Calibration

Calibration Files

750 =PR038492.D 1000 =PR038491.D 500 =PR038493.D
 250 =PR038494.D 50 =PR038495.D

Compound			750	1000	500	250	50	Avg	%RSD	
6)	L1	AR-1016-4	1.882	1.858	1.727	1.734	2.231	1.886	E5	10.87
7)	L1	AR-1016-5	1.828	1.837	1.682	1.724	2.259	1.866	E5	12.31
8)	L2	AR-1221-1			1.266			1.266	E5	0.00
9)	L2	AR-1221-2			9.204			9.204	E4	0.00
10)	L2	AR-1221-3			2.987			2.987	E5	0.00
11)	L3	AR-1232-1			2.131			2.131	E5	0.00
12)	L3	AR-1232-2			1.139			1.139	E5	0.00
13)	L3	AR-1232-3			2.478			2.478	E5	0.00
14)	L3	AR-1232-4			1.209			1.209	E5	0.00
15)	L3	AR-1232-5			8.680			8.680	E4	0.00
16)	L4	AR-1242-1			3.543			3.543	E5	0.00
17)	L4	AR-1242-2			5.091			5.091	E5	0.00
18)	L4	AR-1242-3			3.009			3.009	E5	0.00
19)	L4	AR-1242-4			2.513			2.513	E5	0.00
20)	L4	AR-1242-5			2.887			2.887	E5	0.00
21)	L5	AR-1248-1			2.792			2.792	E5	0.00
22)	L5	AR-1248-2			3.651			3.651	E5	0.00
23)	L5	AR-1248-3			4.247			4.247	E5	0.00
24)	L5	AR-1248-4			5.047			5.047	E5	0.00
25)	L5	AR-1248-5			4.776			4.776	E5	0.00
26)	L6	AR-1254-1			4.652			4.652	E5	0.00
27)	L6	AR-1254-2			6.636			6.636	E5	0.00
28)	L6	AR-1254-3			7.573			7.573	E5	0.00
29)	L6	AR-1254-4			5.763			5.763	E5	0.00
30)	L6	AR-1254-5			6.098			6.098	E5	0.00
31)	L7	AR-1260-1	3.094	3.538	2.919	3.046	4.584	3.436	E5	19.87
32)	L7	AR-1260-2	3.659	4.340	3.502	3.622	5.140	4.053	E5	17.04
33)	L7	AR-1260-3	2.696	3.232	2.703	2.694	3.895	3.044	E5	17.38
34)	L7	AR-1260-4	3.152	3.811	3.168	3.134	4.533	3.560	E5	17.27
35)	L7	AR-1260-5	6.467	8.003	6.579	6.401	8.890	7.268	E5	15.44
36)	L8	AR-1262-1			7.073			7.073	E5	0.00
37)	L8	AR-1262-2			1.416			1.416	E6	0.00
38)	L8	AR-1262-3			9.056			9.056	E5	0.00
39)	L8	AR-1262-4			6.767			6.767	E5	0.00
40)	L8	AR-1262-5			4.918			4.918	E5	0.00
41)	L9	AR-1268-1			1.092			1.092	E6	0.00
42)	L9	AR-1268-2			1.006			1.006	E6	0.00
43)	L9	AR-1268-3			8.215			8.215	E5	0.00
44)	L9	AR-1268-4			3.567			3.567	E5	0.00
45)	L9	AR-1268-5			2.723			2.723	E6	0.00

(#) = Out of Range