

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR060620\
Data File : PR045659.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Jun 2020 13:11
Operator : AJ\MA
Sample : L2899-08DL 5X
Misc :
ALS Vial : 42 Sample Multiplier: 1


Instrument :
ECD_R
ClientSampleId :
ETX78DL

Manual Integrations
APPROVED

Sohil
6/10/2020 8:42:41 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 08 01:13:14 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR060220CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 03 17:48:19 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(21) AR-1248-1 (L5)

R.T.	Response	Conc
5.96	11539938	354.98
6.23	33804745	757.01
6.44	36277124	719.69
6.85	42381607	710.91
6.89	55680677	982.64

(21) AR-1248-1 #2 (L5)

R.T.	Response	Conc
5.09	91364140	362.87
5.33	265276753	853.66
5.37	230919652	710.11
5.55	256111727	626.69
5.95	418223082	1010.79

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR060620\
Data File : PR045659.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Jun 2020 13:11
Operator : AJ\MA
Sample : L2899-08DL 5X
Misc :
ALS Vial : 42 Sample Multiplier: 1


Instrument :
ECD_R
ClientSampleId :
ETX78DL

Manual Integrations
APPROVED

Sohil
6/10/2020 8:42:41 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 08 01:13:14 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR060220CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 03 17:48:19 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(21) AR-1248-1 #2 (L5)
R.T. Response Conc
5.96 11539938 354.98
6.23 33804745 757.01
6.44 36277124 719.69
6.85 42381607 710.91
6.89 55680677 982.64

(21) AR-1248-1 #2 (L5)
R.T. Response Conc
5.09 91364140 362.87
5.33 265276753 853.66
5.37 230919652 710.11
5.55 299786398 733.56
5.95 418223082 1010.79

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR060620\
 Data File : PR045659.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Jun 2020 13:11
 Operator : AJ\MA
 Sample : L2899-08DL 5X
 Misc :
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 ECD_R
 Client Sample ID :
 ETX78DL

Manual Integrations
 APPROVED

Sohil
 6/10/2020 8:42:41 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 08 01:13:14 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR060220CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 03 17:48:19 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

1) SA Tetrachlo...	4.770	3.990	6332301	50972976	3.811	4.440
2) SA Decachlor...	10.787	9.129	8681204	72556335	5.532	6.594

Target Compounds

21) L5 AR-1248-1	5.962	5.092	11539938	91364140	354.981	362.870
22) L5 AR-1248-2	6.234	5.331	33804745	265.3E6	757.005	853.657
23) L5 AR-1248-3	6.443	5.374	36277124	230.9E6	719.691	710.112
24) L5 AR-1248-4	6.852	5.549	42381607	299.8E6	710.915	733.563m
25) L5 AR-1248-5	6.892	5.946	55680677	418.2E6	982.645	1010.788
26) L6 AR-1254-1	6.823	5.904	37196001	466.7E6	578.488	672.611
27) L6 AR-1254-2	7.051	6.052	48973880	164.0E6	481.762	269.089 #
28) L6 AR-1254-3	7.415	6.461	33401975	305.4E6	299.320	298.221
29) L6 AR-1254-4	7.703	6.689	22486610	208.5E6	261.375	311.994
30) L6 AR-1254-5	8.126	7.110	12465258	154.5E6	133.212	173.983 #

ST
 06/11/20

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.