

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR060920CLP.M
 Title : GC EXTRACTABLES
 Last Update : Wed Jun 10 01:45:19 2020
 Response Via : Initial Calibration

Calibration Files

400 =PR0456708.D 1600 =PR0456710.D 800 =PR0456709.D
 200 =PR0456707.D 100 =PR0456706.D

Compound		400	1600	800	200	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.366	1.257	1.418	1.374	1.296	1.342	E6	4.83
2) SA	Decachlorobiphenyl	1.641	1.516	1.564	1.684	1.704	1.622	E6	4.93
3) L1	AR-1016-1	5.158	4.701	4.965	5.316	5.324	5.093	E4	5.16
4) L1	AR-1016-2	7.451	6.811	7.304	7.650	7.490	7.341	E4	4.37
5) L1	AR-1016-3	4.536	4.000	4.322	4.698	4.551	4.421	E4	6.13
6) L1	AR-1016-4	3.771	3.373	3.618	3.888	3.847	3.700	E4	5.66
7) L1	AR-1016-5	3.673	3.298	3.513	3.616	3.725	3.565	E4	4.73
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1	7.301	6.575	6.820	7.411	7.556	7.133	E4	5.84
32) L7	AR-1260-2	9.118	8.305	8.460	9.152	9.458	8.899	E4	5.53
33) L7	AR-1260-3	7.041	6.501	6.634	7.053	7.060	6.858	E4	3.93
34) L7	AR-1260-4	8.384	7.771	7.848	8.367	8.199	8.114	E4	3.55
35) L7	AR-1260-5	1.780	1.719	1.720	1.754	1.775	1.750	E5	1.67
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PR0456708.D 1600 =PR0456710.D 800 =PR0456709.D
 200 =PR0456707.D 100 =PR0456706.D

Compound		400	1600	800	200	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	0.976	0.900	1.020	0.968	0.910	0.955	E7	5.20
2) SA	Decachlorobiphenyl	1.207	1.116	1.156	1.243	1.272	1.199	E7	5.28
3) L1	AR-1016-1	3.981	3.534	3.851	4.032	4.248	3.929	E5	6.70
4) L1	AR-1016-2	4.965	4.810	4.961	5.065	5.148	4.990	E5	2.55
5) L1	AR-1016-3	2.724	2.521	2.686	2.717	2.698	2.669	E5	3.17

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR060920CLP.M
 Title : GC EXTRACTABLES
 Last Update : Wed Jun 10 01:45:19 2020
 Response Via : Initial Calibration

Calibration Files

400 =PR0456708.D 1600 =PR0456710.D 800 =PR0456709.D
 200 =PR0456707.D 100 =PR0456706.D

Compound			400	1600	800	200	100	Avg	%RSD	
6)	L1	AR-1016-4	2.197	1.933	2.094	2.245	2.315	2.157	E5	6.88
7)	L1	AR-1016-5	2.940	2.655	2.834	2.958	3.003	2.878	E5	4.84
8)	L2	AR-1221-1						0.000		-1.00
9)	L2	AR-1221-2						0.000		-1.00
10)	L2	AR-1221-3						0.000		-1.00
11)	L3	AR-1232-1						0.000		-1.00
12)	L3	AR-1232-2						0.000		-1.00
13)	L3	AR-1232-3						0.000		-1.00
14)	L3	AR-1232-4						0.000		-1.00
15)	L3	AR-1232-5						0.000		-1.00
16)	L4	AR-1242-1						0.000		-1.00
17)	L4	AR-1242-2						0.000		-1.00
18)	L4	AR-1242-3						0.000		-1.00
19)	L4	AR-1242-4						0.000		-1.00
20)	L4	AR-1242-5						0.000		-1.00
21)	L5	AR-1248-1						0.000		-1.00
22)	L5	AR-1248-2						0.000		-1.00
23)	L5	AR-1248-3						0.000		-1.00
24)	L5	AR-1248-4						0.000		-1.00
25)	L5	AR-1248-5						0.000		-1.00
26)	L6	AR-1254-1						0.000		-1.00
27)	L6	AR-1254-2						0.000		-1.00
28)	L6	AR-1254-3						0.000		-1.00
29)	L6	AR-1254-4						0.000		-1.00
30)	L6	AR-1254-5						0.000		-1.00
31)	L7	AR-1260-1	5.623	5.211	5.337	5.583	5.759	5.502	E5	4.06
32)	L7	AR-1260-2	7.036	6.488	6.696	7.105	7.168	6.899	E5	4.25
33)	L7	AR-1260-3	6.469	6.168	6.219	6.303	6.367	6.305	E5	1.89
34)	L7	AR-1260-4	5.634	5.371	5.430	5.500	5.589	5.505	E5	1.97
35)	L7	AR-1260-5	1.491	1.441	1.453	1.445	1.442	1.454	E6	1.45
36)	L8	AR-1262-1						0.000		-1.00
37)	L8	AR-1262-2						0.000		-1.00
38)	L8	AR-1262-3						0.000		-1.00
39)	L8	AR-1262-4						0.000		-1.00
40)	L8	AR-1262-5						0.000		-1.00
41)	L9	AR-1268-1						0.000		-1.00
42)	L9	AR-1268-2						0.000		-1.00
43)	L9	AR-1268-3						0.000		-1.00
44)	L9	AR-1268-4						0.000		-1.00
45)	L9	AR-1268-5						0.000		-1.00

 (#) = Out of Range