

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR060920\
 Data File : PR045764.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Jun 2020 12:26
 Operator : AJ\MA
 Sample : L2928-08DL2 20X
 Misc :
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 ETTY1DL2

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 11 16:11:03 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR060920CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 10 18:51:27 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
2)	SA Decachlor...	10.787	9.128	2912709	26299693	1.796	2.193
Target Compounds							
16)	L4 AR-1242-1	5.962	5.093	36982495	294.1E6	692.885	700.111
17)	L4 AR-1242-2	5.986	5.114	65562704	496.9E6	855.590	914.777
18)	L4 AR-1242-3	6.051	5.293	11571638	301.9E6	248.107	1032.116 #
19)	L4 AR-1242-4	6.150	5.375	36611487	231.1E6	954.073	808.618
20)	L4 AR-1242-5	6.891	5.904	22028043	219.5E6	506.582	568.918

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR060920\
Data File : PR045764.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jun 2020 12:26
Operator : AJ\MA
Sample : L2928-08DL2 20X
Misc :
ALS Vial : 56 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
ETTY1DL2

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 11 16:11:03 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR060920CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 10 18:51:27 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

