

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR060920\
 Data File : PR045731.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Jun 2020 21:02
 Operator : AJ\MA
 Sample : AR1254ICC1600
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1254501

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 10 03:34:40 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR060920CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 10 03:32:46 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.773	3.994	127.2E6	968.1E6	72.968	72.205
2) SA Decachlor...	10.786	9.129	265.2E6	2086.5E6	145.690	144.782
Target Compounds						
26) L6 AR-1254-1	6.824	5.906	98117024	1106.1E6	1361.232	1386.681
27) L6 AR-1254-2	7.042	6.054	155.8E6	959.1E6	1381.209	1385.384
28) L6 AR-1254-3	7.416	6.462	175.8E6	1654.3E6	1434.976	1424.112
29) L6 AR-1254-4	7.703	6.691	134.8E6	1078.2E6	1417.237	1419.677
30) L6 AR-1254-5	8.125	7.112	147.3E6	1464.2E6	1435.368	1445.037

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR060920\
Data File : PR045731.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jun 2020 21:02
Operator : AJ\MA
Sample : AR1254ICC1600
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1254501

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 10 03:34:40 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR060920CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 10 03:32:46 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

