

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR062518\
 Data File : PR029332.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jun 2018 22:07
 Operator : UA/SM
 Sample : J3249-05 10X
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 21 06:46:31 2018
 Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_R\Method\PR062518.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 21 03:58:35 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.575	3.722	87336704	217.7E6	4.093	4.989
2)	SA Decachlor...	10.350	8.664	83525295	125.1E6	2.654	2.439
Target Compounds							
6)	L1 AR-1016-4	0.000	5.178	0	177.4E6	N.D.	126.610 #
8)	L2 AR-1221-1	4.752	0.000	14732908	0	50.913	N.D. #
23)	L5 AR-1248-3	0.000	5.178	0	177.4E6	N.D.	91.790 #
25)	L5 AR-1248-5	6.797	0.000	26524638	0	26.078	N.D. #
27)	L6 AR-1254-2	6.797	0.000	26524638	0	15.837	N.D. #
34)	L7 AR-1260-4	8.185	0.000	27610407	0	19.877	N.D. #
39)	L8 AR-1262-4	8.185	0.000	27610407	0	14.676	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR062518\
Data File : PR029332.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jun 2018 22:07
Operator : UA/SM
Sample : J3249-05 10X
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 21 06:46:31 2018
Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_R\Method\PR062518.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 21 03:58:35 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.575 min
Delta R.T.: 0.002 min
Response: 87336704
Conc: 4.09 ng/ml

Instrument :
ECD_R
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.722 min
Delta R.T.: 0.003 min
Response: 217716532
Conc: 4.99 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.350 min
Delta R.T.: 0.000 min
Response: 83525295
Conc: 2.65 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.664 min
Delta R.T.: 0.000 min
Response: 125064385
Conc: 2.44 ng/ml



#6 AR-1016-4


R.T.: 0.000 min
Exp R.T.: 6.208 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#6 AR-1016-4

R.T.: 5.178 min
Delta R.T.: -0.056 min
Response: 177446473
Conc: 126.61 ng/ml




#8 AR-1221-1

R.T.: 4.752 min
Delta R.T.: -0.027 min
Response: 14732908
Conc: 50.91 ng/ml



#8 AR-1221-1


R.T.: 0.000 min
Exp R.T.: 3.936 min
Response: 0
Conc: N.D.



#23 AR-1248-3


R.T.: 0.000 min
Exp R.T.: 6.583 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#23 AR-1248-3

R.T.: 5.178 min
Delta R.T.: -0.056 min
Response: 177446473
Conc: 91.79 ng/ml




#25 AR-1248-5

R.T.: 6.797 min
Delta R.T.: -0.007 min
Response: 26524638
Conc: 26.08 ng/ml



#25 AR-1248-5


R.T.: 0.000 min
Exp R.T.: 5.621 min
Response: 0
Conc: N.D.



#27 AR-1254-2


R.T.: 6.797 min
Delta R.T.: -0.007 min
Response: 26524638
Conc: 15.84 ng/ml

Instrument :
ECD_R
ClientSampleId :




#27 AR-1254-2

R.T.: 0.000 min
Exp R.T. : 5.735 min
Response: 0
Conc: N.D.




#34 AR-1260-4

R.T.: 8.185 min
Delta R.T.: 0.013 min
Response: 27610407
Conc: 19.88 ng/ml



#34 AR-1260-4


R.T.: 0.000 min
Exp R.T. : 7.055 min
Response: 0
Conc: N.D.



#39 AR-1262-4

R.T.: 8.185 min
Delta R.T.: 0.007 min
Response: 27610407
Conc: 14.68 ng/ml

Instrument :
ECD_R
ClientSampleId :



#39 AR-1262-4

R.T.: 0.000 min
Exp R.T. : 7.068 min
Response: 0
Conc: N.D.