

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR062518\
 Data File : PR029301.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jun 2018 13:59
 Operator : UA/SM
 Sample : AR1221ICC500
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1221ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 20 15:08:45 2018
 Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_R\Method\PR062518.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 20 15:08:38 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.574	3.719	1086.1E6	2287.1E6	50.958	52.362
2) SA Decachlor...	10.351	8.665	1540.4E6	2606.9E6	49.172	50.721
Target Compounds						
8) L2 AR-1221-1	4.777	3.930	144.7E6	293.4E6	500.000	500.000
9) L2 AR-1221-2	4.862	4.015	106.2E6	215.0E6	500.000	500.000
10) L2 AR-1221-3	4.939	4.090	336.3E6	691.4E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR062518\
Data File : PR029301.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jun 2018 13:59
Operator : UA/SM
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 20 15:08:45 2018
Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_R\Method\PR062518.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 20 15:08:38 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.574 min
Delta R.T.: 0.000 min
Response: 1086051872
Conc: 50.96 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1221ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.719 min
Delta R.T.: 0.000 min
Response: 2287072569
Conc: 52.36 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.351 min
Delta R.T.: 0.001 min
Response: 1540413235
Conc: 49.17 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.665 min
Delta R.T.: 0.000 min
Response: 2606892979
Conc: 50.72 ng/ml



#8 AR-1221-1


R.T.: 4.777 min
Delta R.T.: -0.002 min
Response: 144686095
Conc: 500.00 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1221ICC500




#8 AR-1221-1

R.T.: 3.930 min
Delta R.T.: -0.007 min
Response: 293361263
Conc: 500.00 ng/ml




#9 AR-1221-2

R.T.: 4.862 min
Delta R.T.: -0.002 min
Response: 106206510
Conc: 500.00 ng/ml



#9 AR-1221-2


R.T.: 4.015 min
Delta R.T.: -0.006 min
Response: 214993011
Conc: 500.00 ng/ml



#10 AR-1221-3

R.T.: 4.939 min
Delta R.T.: -0.002 min
Response: 336310532
Conc: 500.00 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1221ICC500



#10 AR-1221-3

R.T.: 4.090 min
Delta R.T.: -0.007 min
Response: 691421442
Conc: 500.00 ng/ml