

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR062518\
 Data File : PR029347.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Jun 2018 01:56
 Operator : UA/SM
 Sample : J3523-01 10X
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AREA1(S)

Manual Integrations
 APPROVED

Sohil
 6/23/2018 1:43:08 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 21 07:34:21 2018
 Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_R\Method\PR062518.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 21 03:58:35 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.573	3.722	23682890	50593706	1.110	1.159
2) SA Decachlor...	10.346	8.661	35973640	39477958	1.143	0.770m#

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR062518\
Data File : PR029347.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jun 2018 01:56
Operator : UA/SM
Sample : J3523-01 10X
Misc :
ALS Vial : 25 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
AREA1(S)


Manual Integrations
APPROVED

Sohil
6/23/2018 1:43:08 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 21 07:34:21 2018
Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_R\Method\PR062518.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 21 03:58:35 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.573 min
Delta R.T.: 0.000 min
Response: 23682890
Conc: 1.11 ng/ml

Instrument :
ECD_R
ClientSampleId :
AREA1(S)


Manual Integrations
APPROVED

Sohil
6/23/2018 1:43:08 AM




#1 Tetrachloro-m-xylene

R.T.: 3.722 min
Delta R.T.: 0.004 min
Response: 50593706
Conc: 1.16 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.346 min
Delta R.T.: -0.004 min
Response: 35973640
Conc: 1.14 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.661 min
Delta R.T.: -0.003 min
Response: 39477958
Conc: 0.77 ng/ml m