

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR062618\
 Data File : PR029358.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Jun 2018 18:26
 Operator : UA/SM
 Sample : J3529-02RE
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 TPTW-03RE

Manual Integrations
 APPROVED

ugo
 6/25/2018 2:55:06 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 21 18:39:58 2018
 Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_R\Method\PR062518.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 21 03:58:35 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.571	3.721	299.8E6	910.9E6	14.049m	20.871m#
2) SA Decachlor...	10.339	8.660	153.2E6	130.8E6	4.866m	2.552 #

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR062618\
Data File : PR029358.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jun 2018 18:26
Operator : UA/SM
Sample : J3529-02RE
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
TPTW-03RE

Manual Integrations
APPROVED

ugo
6/25/2018 2:55:06 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 21 18:39:58 2018
Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_R\Method\PR062518.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 21 03:58:35 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

