

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR071119.M
 Title : GC EXTRACTABLES
 Last Update : Fri Jul 12 04:20:28 2019
 Response Via : Initial Calibration

Calibration Files

750 =PR039216.D 1000 =PR039215.D 500 =PR039217.D
 250 =PR039218.D 50 =PR039219.D

Compound		750	1000	500	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	2.538	2.448	2.905	2.728	3.030	2.730	E6	8.91
2) SA	Decachlorobiphenyl	1.990	1.884	1.989	1.981	2.603	2.090	E6	13.91
3) L1	AR-1016-1	0.776	0.733	0.913	0.884	1.149	0.891	E5	18.22
4) L1	AR-1016-2	1.279	1.212	1.494	1.422	1.705	1.422	E5	13.62
5) L1	AR-1016-3	6.087	5.757	7.092	6.852	8.209	6.799	E4	14.08
6) L1	AR-1016-4	4.779	4.441	5.596	5.527	6.672	5.403	E4	15.97
7) L1	AR-1016-5	6.119	5.792	7.227	7.070	8.403	6.922	E4	14.85
8) L2	AR-1221-1			2.364			2.364	E4	0.00
9) L2	AR-1221-2			1.795			1.795	E4	0.00
10) L2	AR-1221-3			5.664			5.664	E4	0.00
11) L3	AR-1232-1			4.897			4.897	E4	0.00
12) L3	AR-1232-2			5.298			5.298	E4	0.00
13) L3	AR-1232-3			2.435			2.435	E4	0.00
14) L3	AR-1232-4			2.114			2.114	E4	0.00
15) L3	AR-1232-5			2.352			2.352	E4	0.00
16) L4	AR-1242-1			7.257			7.257	E4	0.00
17) L4	AR-1242-2			1.178			1.178	E5	0.00
18) L4	AR-1242-3			5.713			5.713	E4	0.00
19) L4	AR-1242-4			5.445			5.445	E4	0.00
20) L4	AR-1242-5			7.262			7.262	E4	0.00
21) L5	AR-1248-1			5.524			5.524	E4	0.00
22) L5	AR-1248-2			7.583			7.583	E4	0.00
23) L5	AR-1248-3			7.749			7.749	E4	0.00
24) L5	AR-1248-4			9.430			9.430	E4	0.00
25) L5	AR-1248-5			9.594			9.594	E4	0.00
26) L6	AR-1254-1			1.187			1.187	E5	0.00
27) L6	AR-1254-2			1.010			1.010	E5	0.00
28) L6	AR-1254-3			1.682			1.682	E5	0.00
29) L6	AR-1254-4			1.068			1.068	E5	0.00
30) L6	AR-1254-5			1.444			1.444	E5	0.00
31) L7	AR-1260-1	1.077	1.012	1.177	1.184	1.473	1.185	E5	14.88
32) L7	AR-1260-2	1.328	1.272	1.412	1.447	1.876	1.467	E5	16.26
33) L7	AR-1260-3	1.245	1.179	1.305	1.332	1.648	1.342	E5	13.47
34) L7	AR-1260-4	1.004	0.958	1.028	1.048	1.243	1.056	E5	10.40
35) L7	AR-1260-5	2.604	2.506	2.599	2.617	3.011	2.668	E5	7.39
36) L8	AR-1262-1			7.095			7.095	E4	0.00
37) L8	AR-1262-2			3.402			3.402	E5	0.00
38) L8	AR-1262-3			1.304			1.304	E5	0.00
39) L8	AR-1262-4			2.475			2.475	E5	0.00
40) L8	AR-1262-5			1.142			1.142	E5	0.00
41) L9	AR-1268-1			3.440			3.440	E5	0.00
42) L9	AR-1268-2			3.188			3.188	E5	0.00
43) L9	AR-1268-3			2.581			2.581	E5	0.00
44) L9	AR-1268-4			1.124			1.124	E5	0.00
45) L9	AR-1268-5			8.694			8.694	E5	0.00

Signal #2 Calibration Files

750 =PR039216.D 1000 =PR039215.D 500 =PR039217.D
 250 =PR039218.D 50 =PR039219.D

Compound		750	1000	500	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	0.878	0.837	1.023	0.975	1.111	0.965	E7	11.43
2) SA	Decachlorobiphenyl	6.515	6.374	6.499	6.695	8.132	6.843	E6	10.66
3) L1	AR-1016-1	2.717	2.576	3.207	3.151	3.836	3.097	E5	15.96
4) L1	AR-1016-2	3.919	3.714	4.618	4.441	5.370	4.412	E5	14.73
5) L1	AR-1016-3	2.247	2.129	2.663	2.617	3.126	2.557	E5	15.38

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR071119.M
 Title : GC EXTRACTABLES
 Last Update : Fri Jul 12 04:20:28 2019
 Response Via : Initial Calibration

Calibration Files

750 =PR039216.D 1000 =PR039215.D 500 =PR039217.D
 250 =PR039218.D 50 =PR039219.D

Compound			750	1000	500	250	50	Avg	%RSD	
6)	L1	AR-1016-4	1.872	1.799	2.231	2.181	2.643	2.145	E5	15.65
7)	L1	AR-1016-5	1.781	1.680	2.067	2.023	2.475	2.005	E5	15.39
8)	L2	AR-1221-1			7.976			7.976	E4	0.00
9)	L2	AR-1221-2			5.924			5.924	E4	0.00
10)	L2	AR-1221-3			1.922			1.922	E5	0.00
11)	L3	AR-1232-1			1.626			1.626	E5	0.00
12)	L3	AR-1232-2			7.836			7.836	E4	0.00
13)	L3	AR-1232-3			1.582			1.582	E5	0.00
14)	L3	AR-1232-4			7.762			7.762	E4	0.00
15)	L3	AR-1232-5			5.486			5.486	E4	0.00
16)	L4	AR-1242-1			2.584			2.584	E5	0.00
17)	L4	AR-1242-2			3.653			3.653	E5	0.00
18)	L4	AR-1242-3			2.149			2.149	E5	0.00
19)	L4	AR-1242-4			1.793			1.793	E5	0.00
20)	L4	AR-1242-5			1.876			1.876	E5	0.00
21)	L5	AR-1248-1			1.911			1.911	E5	0.00
22)	L5	AR-1248-2			2.447			2.447	E5	0.00
23)	L5	AR-1248-3			2.778			2.778	E5	0.00
24)	L5	AR-1248-4			3.242			3.242	E5	0.00
25)	L5	AR-1248-5			3.034			3.034	E5	0.00
26)	L6	AR-1254-1			2.531			2.531	E5	0.00
27)	L6	AR-1254-2			3.855			3.855	E5	0.00
28)	L6	AR-1254-3			4.212			4.212	E5	0.00
29)	L6	AR-1254-4			3.129			3.129	E5	0.00
30)	L6	AR-1254-5			3.297			3.297	E5	0.00
31)	L7	AR-1260-1	2.971	2.797	3.232	3.235	4.109	3.269	E5	15.44
32)	L7	AR-1260-2	3.640	3.476	3.856	3.902	4.733	3.921	E5	12.37
33)	L7	AR-1260-3	2.764	2.663	2.869	2.890	3.532	2.943	E5	11.59
34)	L7	AR-1260-4	3.226	3.112	3.299	3.309	3.964	3.382	E5	9.90
35)	L7	AR-1260-5	6.931	6.781	6.937	6.949	8.097	7.139	E5	7.56
36)	L8	AR-1262-1			4.863			4.863	E5	0.00
37)	L8	AR-1262-2			9.290			9.290	E5	0.00
38)	L8	AR-1262-3			6.058			6.058	E5	0.00
39)	L8	AR-1262-4			4.673			4.673	E5	0.00
40)	L8	AR-1262-5			3.390			3.390	E5	0.00
41)	L9	AR-1268-1			9.634			9.634	E5	0.00
42)	L9	AR-1268-2			9.114			9.114	E5	0.00
43)	L9	AR-1268-3			7.614			7.614	E5	0.00
44)	L9	AR-1268-4			3.327			3.327	E5	0.00
45)	L9	AR-1268-5			2.652			2.652	E6	0.00

 (#) = Out of Range