

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR071119\
 Data File : PR039240.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jul 2019 20:19
 Operator : SM\AJ
 Sample : K3730-01
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 13773

Manual Integrations
 APPROVED

Ankita
 7/12/2019 1:31:14 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 12 07:21:39 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR071119.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jul 12 04:20:28 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

1) SA Tetrachlo...	3.761	4.608	73536124	272.0E6	26.938m	28.194m
2) SA Decachlor...	8.700	10.343	70152145	237.4E6	33.572	34.700

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR071119\
Data File : PR039240.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jul 2019 20:19
Operator : SM\AJ
Sample : K3730-01
Misc :
ALS Vial : 26 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
13773


Manual Integrations
APPROVED

Ankita
7/12/2019 1:31:14 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 12 07:21:39 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR071119.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jul 12 04:20:28 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.761 min
Delta R.T.: 0.002 min
Response: 73536124
Conc: 26.94 ng/ml m

Instrument :
ECD_R
ClientSampleId :
13773


Manual Integrations
APPROVED

Ankita
7/12/2019 1:31:14 PM




#1 Tetrachloro-m-xylene

R.T.: 4.608 min
Delta R.T.: 0.002 min
Response: 272020544
Conc: 28.19 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.700 min
Delta R.T.: -0.003 min
Response: 70152145
Conc: 33.57 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.343 min
Delta R.T.: -0.001 min
Response: 237448190
Conc: 34.70 ng/ml