

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR071819\
 Data File : PR039402.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Jul 2019 18:33
 Operator : SM\AJ
 Sample : K3852-01
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 COMP-1_071619

Manual Integrations
 APPROVED

Ankita
 7/21/2019 11:52:32 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 19 01:34:19 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR071519.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jul 16 02:56:01 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.758	4.605	50602712	206.5E6	32.670	31.927
2) SA Decachlor...	8.696	10.338	35516044	151.2E6	14.524	17.667m
Target Compounds						
26) L6 AR-1254-1	5.613	6.604	3161493	9150556	23.904m	28.578m
27) L6 AR-1254-2	5.757	6.817	2552230	17638010	21.788	34.468m#
28) L6 AR-1254-3	6.153	7.184	8278885	6100422	40.066m	10.166 #
29) L6 AR-1254-4	6.377	7.464	3266816	13123426	21.339m	30.324 #
30) L6 AR-1254-5	6.790	7.879	6188391	17645496	33.218m	37.546

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR071819\
Data File : PR039402.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Jul 2019 18:33
Operator : SM\AJ
Sample : K3852-01
Misc :
ALS Vial : 8 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
COMP-1_071619


Manual Integrations
APPROVED

Ankita
7/21/2019 11:52:32 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 19 01:34:19 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR071519.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jul 16 02:56:01 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.758 min
Delta R.T.: -0.005 min
Response: 50602712
Conc: 32.67 ng/ml

Instrument :
ECD_R
ClientSampleId :
COMP-1_071619


Manual Integrations
APPROVED

Ankita
7/21/2019 11:52:32 AM




#1 Tetrachloro-m-xylene

R.T.: 4.605 min
Delta R.T.: -0.005 min
Response: 206481630
Conc: 31.93 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.696 min
Delta R.T.: -0.010 min
Response: 35516044
Conc: 14.52 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.338 min
Delta R.T.: -0.011 min
Response: 151186654
Conc: 17.67 ng/ml m




#26 AR-1254-1

R.T.: 5.613 min
Delta R.T.: -0.005 min
Response: 3161493
Conc: 23.90 ng/ml m

Instrument :
ECD_R
ClientSampleId :
COMP-1_071619


Manual Integrations
APPROVED

Ankita
7/21/2019 11:52:32 AM




#26 AR-1254-1

R.T.: 6.604 min
Delta R.T.: -0.005 min
Response: 9150556
Conc: 28.58 ng/ml m




#27 AR-1254-2

R.T.: 5.757 min
Delta R.T.: -0.006 min
Response: 2552230
Conc: 21.79 ng/ml



#27 AR-1254-2

R.T.: 6.817 min
Delta R.T.: -0.006 min
Response: 17638010
Conc: 34.47 ng/ml m




#28 AR-1254-3

R.T.: 6.153 min
Delta R.T.: -0.007 min
Response: 8278885
Conc: 40.07 ng/ml m

Instrument :
ECD_R
ClientSampleId :
COMP-1_071619


Manual Integrations
APPROVED

Ankita
7/21/2019 11:52:32 AM




#28 AR-1254-3

R.T.: 7.184 min
Delta R.T.: -0.005 min
Response: 6100422
Conc: 10.17 ng/ml




#29 AR-1254-4

R.T.: 6.377 min
Delta R.T.: -0.008 min
Response: 3266816
Conc: 21.34 ng/ml m



#29 AR-1254-4

R.T.: 7.464 min
Delta R.T.: -0.006 min
Response: 13123426
Conc: 30.32 ng/ml




#30 AR-1254-5

R.T.: 6.790 min
Delta R.T.: -0.007 min
Response: 6188391
Conc: 33.22 ng/ml m

Instrument :
ECD_R
ClientSampleId :
COMP-1_071619

Manual Integrations
APPROVED

Ankita
7/21/2019 11:52:32 AM



#30 AR-1254-5

R.T.: 7.879 min
Delta R.T.: -0.006 min
Response: 17645496
Conc: 37.55 ng/ml