Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR072219\
Data File : PR039577.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Jul 2019 16:53
Operator : SM\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 68 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
ICVPR072219AR1254

Manual Integrations APPROVED
Ankita
7/23/2019 5:12:49 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 22 17:17:26 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR072219.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jul 22 14:03:19 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	3.759	4.600	129.9E6	515.5E6	54.106	51.427
2)	SA Decachlor...	8.704	10.341	111.2E6	403.4E6	44.759	42.970
	Target Compounds						
26)	L6 AR-1254-1	5.617	6.602	77175567	182.7E6	500.392	501.677
27)	L6 AR-1254-2	5.761	6.816	65323722	278.0E6	498.910	486.688
28)	L6 AR-1254-3	6.159	7.182	109.0E6	297.6E6	487.750	477.280
29)	L6 AR-1254-4	6.383	7.464	70382931	219.8E6	484.006m	485.274
30)	L6 AR-1254-5	6.795	7.879	92320975	222.5E6	484.839	468.516

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR072219\
Data File : PR039577.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Jul 2019 16:53
Operator : SM\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 68 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
ICVPR072219AR1254

Manual Integrations APPROVED
Ankita
7/23/2019 5:12:49 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 22 17:17:26 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR072219.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jul 22 14:03:19 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm