

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR080218CLP.M
 Title : GC EXTRACTABLES
 Last Update : Thu Aug 02 01:57:59 2018
 Response Via : Initial Calibration

Calibration Files

400 =PR031032.D 1600 =PR031034.D 800 =PR031033.D
 200 =PR031031.D 100 =PR031030.D

Compound		400	1600	800	200	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.298	1.242	1.190	1.415	1.479	1.325	E7	9.10
2) SA	Decachlorobiphenyl	3.285	2.916	2.986	3.577	3.793	3.312	E7	11.35
3) L1	AR-1016-1	3.668	3.308	3.309	4.109	4.312	3.741	E5	12.26
4) L1	AR-1016-2	3.683	3.369	3.346	3.995	4.142	3.707	E5	9.70
5) L1	AR-1016-3	6.057	5.348	5.349	6.604	7.010	6.073	E5	12.24
6) L1	AR-1016-4	5.320	4.761	4.807	5.823	6.017	5.346	E5	10.71
7) L1	AR-1016-5	4.700	4.141	4.194	5.271	5.266	4.714	E5	11.69
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1	1.145	1.021	1.030	1.270	1.340	1.161	E6	12.24
32) L7	AR-1260-2	1.674	1.500	1.514	1.845	1.908	1.688	E6	11.03
33) L7	AR-1260-3	1.901	1.721	1.745	2.097	2.154	1.924	E6	10.29
34) L7	AR-1260-4	1.407	1.258	1.269	1.528	1.544	1.401	E6	9.73
35) L7	AR-1260-5	3.476	3.179	3.197	3.768	3.878	3.500	E6	9.15
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PR031032.D 1600 =PR031034.D 800 =PR031033.D
 200 =PR031031.D 100 =PR031030.D

Compound		400	1600	800	200	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	2.150	2.045	1.948	2.368	2.515	2.205	E7	10.57
2) SA	Decachlorobiphenyl	1.967	1.696	1.751	2.186	2.338	1.988	E7	13.85
3) L1	AR-1016-1	6.802	5.951	6.059	7.717	8.258	6.957	E5	14.57
4) L1	AR-1016-2	1.071	0.945	0.935	1.187	1.321	1.092	E6	15.06
5) L1	AR-1016-3	1.782	1.548	1.555	2.041	2.109	1.807	E6	14.57

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR080218CLP.M
 Title : GC EXTRACTABLES
 Last Update : Thu Aug 02 01:57:59 2018
 Response Via : Initial Calibration

Calibration Files

400 =PR031032.D 1600 =PR031034.D 800 =PR031033.D
 200 =PR031031.D 100 =PR031030.D

Compound			400	1600	800	200	100	Avg	%RSD

6)	L1	AR-1016-4	1.170	1.025	1.043	1.340	1.468	1.209 E6	15.85
7)	L1	AR-1016-5	1.082	0.942	0.959	1.212	1.295	1.098 E6	14.08
8)	L2	AR-1221-1						0.000	-1.00
9)	L2	AR-1221-2						0.000	-1.00
10)	L2	AR-1221-3						0.000	-1.00
11)	L3	AR-1232-1						0.000	-1.00
12)	L3	AR-1232-2						0.000	-1.00
13)	L3	AR-1232-3						0.000	-1.00
14)	L3	AR-1232-4						0.000	-1.00
15)	L3	AR-1232-5						0.000	-1.00
16)	L4	AR-1242-1						0.000	-1.00
17)	L4	AR-1242-2						0.000	-1.00
18)	L4	AR-1242-3						0.000	-1.00
19)	L4	AR-1242-4						0.000	-1.00
20)	L4	AR-1242-5						0.000	-1.00
21)	L5	AR-1248-1						0.000	-1.00
22)	L5	AR-1248-2						0.000	-1.00
23)	L5	AR-1248-3						0.000	-1.00
24)	L5	AR-1248-4						0.000	-1.00
25)	L5	AR-1248-5						0.000	-1.00
26)	L6	AR-1254-1						0.000	-1.00
27)	L6	AR-1254-2						0.000	-1.00
28)	L6	AR-1254-3						0.000	-1.00
29)	L6	AR-1254-4						0.000	-1.00
30)	L6	AR-1254-5						0.000	-1.00
31)	L7	AR-1260-1	2.582	2.226	2.289	2.901	3.178	2.635 E6	15.36
32)	L7	AR-1260-2	3.211	2.780	2.861	3.595	3.931	3.276 E6	14.90
33)	L7	AR-1260-3	2.677	2.320	2.391	2.998	3.214	2.720 E6	14.13
34)	L7	AR-1260-4	1.716	1.424	1.505	1.927	2.071	1.729 E6	15.82
35)	L7	AR-1260-5	3.656	3.050	3.182	4.136	4.514	3.708 E6	16.76
36)	L8	AR-1262-1						0.000	-1.00
37)	L8	AR-1262-2						0.000	-1.00
38)	L8	AR-1262-3						0.000	-1.00
39)	L8	AR-1262-4						0.000	-1.00
40)	L8	AR-1262-5						0.000	-1.00
41)	L9	AR-1268-1						0.000	-1.00
42)	L9	AR-1268-2						0.000	-1.00
43)	L9	AR-1268-3						0.000	-1.00
44)	L9	AR-1268-4						0.000	-1.00
45)	L9	AR-1268-5						0.000	-1.00

 (#) = Out of Range