

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR080218CLP.M
 Title : GC EXTRACTABLES
 Last Update : Thu Aug 02 02:33:30 2018
 Response Via : Initial Calibration

Calibration Files

400 =PR031039.D 1600 =PR031041.D 800 =PR031040.D
 200 =PR031038.D 100 =PR031037.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.305	1.322	1.285	1.429	1.353	1.339	E7	4.19
2) SA	Decachlorobiphenyl	3.287	2.934	3.064	3.612	3.901	3.360	E7	11.82
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1	1.733	1.636	1.649	1.970	2.044	1.806	E5	10.45
17) L4	AR-1242-2	3.373	3.079	3.211	3.693	3.927	3.457	E5	10.10
18) L4	AR-1242-3	1.704	1.581	1.645	2.069	2.220	1.844	E5	15.35
19) L4	AR-1242-4	1.202	1.184	1.195	1.400	1.383	1.273	E5	8.54
20) L4	AR-1242-5	4.394	4.190	4.261	4.856	4.938	4.528	E5	7.65
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PR031039.D 1600 =PR031041.D 800 =PR031040.D
 200 =PR031038.D 100 =PR031037.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.029	2.009	1.971	2.208	2.120	2.067	E7	4.63
2) SA	Decachlorobiphenyl	2.002	1.715	1.830	2.183	2.511	2.048	E7	15.30
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR080218CLP.M
 Title : GC EXTRACTABLES
 Last Update : Thu Aug 02 02:33:30 2018
 Response Via : Initial Calibration

Calibration Files

400 =PR031039.D 1600 =PR031041.D 800 =PR031040.D
 200 =PR031038.D 100 =PR031037.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1	3.769	3.350	3.468	4.315	3.962	3.773 E5	10.28
17) L4	AR-1242-2	7.615	7.046	7.287	8.322	8.145	7.683 E5	7.10
18) L4	AR-1242-3	6.458	5.596	5.964	7.380	7.534	6.586 E5	12.96
19) L4	AR-1242-4	0.980	0.847	0.911	1.131	1.117	0.997 E6	12.54
20) L4	AR-1242-5	6.900	5.932	6.400	7.653	7.038	6.785 E5	9.63
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range